

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS092718\
 Data File : PS001836.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Sep 2018 17:55
 Operator : EI\SJ
 Sample : J5099-02
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 28 01:59:54 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS091718.M
 Quant Title : 8080.M
 QLast Update : Tue Sep 18 03:01:16 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.457	10.953	77882552	77122042	347.201	303.610
Target Compounds							
1) T	Dalapon	2.888	2.369	-95146	45844594	N.D.	155.186
2) T	3,5-DICHL...	0.000	9.193	0	23262611	N.D.	73.701 #
6) T	MCPD	12.144	0.000	28897090	0	45.348	N.D. #
15) T	Picloram	0.000	16.890	0	14373772	N.D.	8.451 #
16) T	DCPA	0.000	16.890	0	14373772	N.D.	7.793 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS092718\
Data File : PS001836.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Sep 2018 17:55
Operator : EI\SJ
Sample : J5099-02
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 28 01:59:54 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS091718.M
Quant Title : 8080.M
QLast Update : Tue Sep 18 03:01:16 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 2.888 min
Delta R.T.: -0.061 min
Response: -95146
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :




#1 Dalapon

R.T.: 2.369 min
Delta R.T.: -0.041 min
Response: 45844594
Conc: 155.19 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 0.000 min
Exp R.T. : 9.902 min
Response: 0
Conc: N.D.



#2 3,5-DICHLOROBENZOIC ACID


R.T.: 9.193 min
Delta R.T.: 0.067 min
Response: 23262611
Conc: 73.70 ng/ml



#4 2,4-DCAA


R.T.: 11.457 min
Delta R.T.: -0.023 min
Response: 77882552
Conc: 347.20 ng/ml

Instrument :
ECD_S
ClientSampleId :




#4 2,4-DCAA

R.T.: 10.953 min
Delta R.T.: -0.023 min
Response: 77122042
Conc: 303.61 ng/ml




#6 MCPP

R.T.: 12.144 min
Delta R.T.: 0.019 min
Response: 28897090
Conc: 45.35 ug/ml



#6 MCPP


R.T.: 0.000 min
Exp R.T. : 11.483 min
Response: 0
Conc: N.D.



#15 Picloram


R.T.: 0.000 min
Exp R.T. : 16.988 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :




#15 Picloram

R.T.: 16.890 min
Delta R.T.: 0.040 min
Response: 14373772
Conc: 8.45 ng/ml



#16 DCPA

R.T.: 0.000 min
Exp R.T. : 17.565 min
Response: 0
Conc: N.D.



#16 DCPA

R.T.: 16.890 min
Delta R.T.: -0.016 min
Response: 14373772
Conc: 7.79 ng/ml