

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS092719\
 Data File : PS006933.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Sep 2019 12:21
 Operator : SM\AJ
 Sample : K4657-03MS
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 BU-02-092419MS

Manual Integrations
 APPROVED

Ankita
 9/30/2019 10:05:25 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 28 01:10:01 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS092019.M
 Quant Title : 8080.M
 QLast Update : Fri Sep 20 11:14:00 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.486	5.217	550.2E6	235.6E6	602.144	512.251m
Target Compounds							
1) T	Dalapon	2.053	1.763	393.4E6	232.0E6	368.215	435.277
2) T	3,5-DICHL...	4.863	4.520	1360.2E6	182.6E6	1250.469	317.625m#
3) T	4-Nitroph...	5.342	4.905	47858795	41701392	125.516m	172.512m#
5) T	DICAMBA	5.638	5.350	1207.1E6	998.8E6	373.597	593.344 #
6) T	MCPD	5.787	5.428	163.0E6	69564913	76.617	59.891
7) T	MCPA	5.921	5.603	164.8E6	104.6E6	52.977	56.808m
8) T	DICHLORPROP	6.266	5.884	234.2E6	153.5E6	305.054m	370.637
9) T	2,4-D	6.492	6.155	2574.3E6	292.0E6	3141.246m	572.456 #
10) T	Pentachlo...	6.743	6.522	4013.3E6	1983.8E6	336.737	315.320
11) T	2,4,5-TP ...	7.276	6.884	1525.1E6	741.8E6	335.685	309.694
12) T	2,4,5-T	7.549	7.243	1704.2E6	712.1E6	334.917	329.442m
13) T	2,4-DB	8.085	7.732	1014.4E6	112.1E6	1748.956	394.044 #
14) T	DINOSEB	9.167	8.043	491.4E6	333.8E6	129.392	193.784 #
15) T	Picloram	9.004	8.978	2523.5E6	1123.9E6	376.317	308.800
16) T	DCPA	9.440	8.934	1474.0E6	586.1E6	252.596	184.467 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS092719\
Data File : PS006933.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Sep 2019 12:21
Operator : SM\AJ
Sample : K4657-03MS
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
BU-02-092419MS

Manual Integrations
APPROVED

Ankita
9/30/2019 10:05:25 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 28 01:10:01 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS092019.M
Quant Title : 8080.M
QLast Update : Fri Sep 20 11:14:00 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

