

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS092719\
 Data File : PS006934.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Sep 2019 12:43
 Operator : SM\AJ
 Sample : K4657-03MSD
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 BU-02-092419MSD

Manual Integrations
 APPROVED

Ankita
 9/30/2019 10:05:27 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 28 01:10:20 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS092019.M
 Quant Title : 8080.M
 QLast Update : Fri Sep 20 11:14:00 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.478	5.209	552.7E6	229.0E6	604.825	497.927m
Target Compounds							
1) T	Dalapon	2.045	1.755	373.9E6	255.4E6	349.957	479.126m#
2) T	3,5-DICHL...	4.856	4.513	1413.0E6	183.8E6	1299.083	319.703m#
3) T	4-Nitroph...	5.335	4.899	57889143	52982458	151.821m	219.180 #
5) T	DICAMBA	5.630	5.341	1208.0E6	947.3E6	373.883	562.769m#
6) T	MCPD	5.779	5.421	118.7E6	52121102	55.816m	44.873m
7) T	MCPA	5.913	5.596	113.3E6	101.5E6	36.430m	55.158m#
8) T	DICHLORPROP	6.259	5.877	227.3E6	150.8E6	296.075m	364.222
9) T	2,4-D	6.484	6.148	2666.6E6	299.6E6	3253.937m	587.399 #
10) T	Pentachlo...	6.735	6.514	3831.6E6	1993.0E6	321.492m	316.776m
11) T	2,4,5-TP ...	7.268	6.876	1524.5E6	742.1E6	335.539	309.823
12) T	2,4,5-T	7.541	7.236	1708.5E6	699.9E6	335.761	323.799m
13) T	2,4-DB	8.076	7.724	1052.9E6	108.5E6	1815.297	381.471 #
14) T	DINOSEB	9.160	8.037	496.5E6	330.3E6	130.719	191.753 #
15) T	Picloram	8.995	8.971	2532.9E6	1133.4E6	377.722	311.393
16) T	DCPA	9.432	8.927	1408.7E6	585.7E6	241.404	184.324m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS092719\
Data File : PS006934.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Sep 2019 12:43
Operator : SM\AJ
Sample : K4657-03MSD
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
BU-02-092419MSD

Manual Integrations
APPROVED

Ankita
9/30/2019 10:05:27 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 28 01:10:20 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS092019.M
Quant Title : 8080.M
QLast Update : Fri Sep 20 11:14:00 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

