

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS092721\
 Data File : PS016901.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Sep 2021 09:49
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 28 01:37:33 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS092021.M
 Quant Title : 8080.M
 QLast Update : Tue Sep 21 01:52:13 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.159	6.749	1102.7E6	428.5E6	723.762	730.695
Target Compounds							
1) T	Dalapon	2.998	2.500	1585.0E6	494.5E6	664.705	680.390
2) T	3,5-DICHL...	6.382	5.944	1417.2E6	604.2E6	665.954	685.268
3) T	4-Nitroph...	6.989	6.373	566.6E6	206.7E6	654.122	661.005
5) T	DICAMBA	7.349	6.911	4080.6E6	1796.0E6	699.169	705.014
6) T	MCPD	7.517	7.007	267.8E6	144.1E6	61.816	70.348
7) T	MCPA	7.665	7.218	432.9E6	200.7E6	69.990	65.914
8) T	DICHLORPROP	8.038	7.527	1016.1E6	445.4E6	688.183	686.736
9) T	2,4-D	8.266	7.817	1225.3E6	508.7E6	696.887	668.671
10) T	Pentachlo...	8.572	8.243	14180.6E6	5784.7E6	701.787	702.964
11) T	2,4,5-TP ...	9.120	8.602	5584.8E6	2462.3E6	699.383	703.291
12) T	2,4,5-T	9.404	8.979	5762.0E6	2267.4E6	698.129	696.742
13) T	2,4-DB	9.952	9.484	1148.4E6	285.7E6	694.738	685.698
14) T	DINOSEB	11.104	9.809	4229.3E6	1753.0E6	694.178	715.651
15) T	Picloram	10.926	10.786	8256.4E6	2890.6E6	693.706	673.842
16) T	DCPA	11.372	10.737	7094.0E6	3076.6E6	690.898	705.819m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS092721\
Data File : PS016901.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Sep 2021 09:49
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 28 01:37:33 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS092021.M
Quant Title : 8080.M
QLast Update : Tue Sep 21 01:52:13 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

