

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS092723\
 Data File : PS024398.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Sep 2023 17:45
 Operator : AR\AJ
 Sample : PB155875BS
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 PB155875BS

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 09/28/2023
 Supervised By : Ankita Jodhani 09/28/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 28 00:35:34 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS082823.M
 Quant Title : 8080.M
 QLast Update : Mon Aug 28 16:48:10 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.361	6.786	1011.2E6	335.0E6	507.557	498.817
Target Compounds							
1) T	Dalapon	2.482	2.547	959.6E6	447.1E6	441.484	423.951
2) T	3,5-DICHL...	5.736	5.971	1445.5E6	484.1E6	475.544	464.316
3) T	4-Nitroph...	6.199	6.399	561.4E6	174.9E6	451.843	414.716
5) T	DICAMBA	6.503	6.948	4206.1E6	1362.4E6	482.503	473.030
6) T	MCPD	6.653	7.043	267.2E6	90166673	42.901	47.068m
7) T	MCPA	6.776	7.249	396.4E6	129.1E6	45.669	44.181m
8) T	DICHLORPROP	7.091	7.563	1064.4E6	311.4E6	468.201	428.597
9) T	2,4-D	7.288	7.847	1232.8E6	404.1E6	463.494	453.605
10) T	Pentachlo...	7.517	8.271	15973.0E6	5189.5E6	510.766	489.819
11) T	2,4,5-TP ...	8.039	8.630	6518.9E6	1959.6E6	487.698	472.278
12) T	2,4,5-T	8.293	9.000	5954.0E6	1852.8E6	480.550	461.692
13) T	2,4-DB	8.798	9.503	947.1E6	230.7E6	439.351	501.991
14) T	DINOSEB	9.851	9.831	4641.6E6	1378.0E6	480.373	460.975
15) T	Picloram	9.677	10.783	6878.7E6	2442.9E6	421.202	411.346
16) T	DCPA	10.117	10.754	8044.1E6	2520.3E6	496.736	491.292

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS092723\
Data File : PS024398.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Sep 2023 17:45
Operator : AR\AJ
Sample : PB155875BS
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
PB155875BS

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 09/28/2023
Supervised By : Ankita Jodhani 09/28/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 28 00:35:34 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS082823.M
Quant Title : 8080.M
QLast Update : Mon Aug 28 16:48:10 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

