

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS092821\
 Data File : PS016917.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Sep 2021 17:49
 Operator : DD\AJ
 Sample : M3870-01MS
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 TP-CMS

Manual Integrations
 APPROVED

mohammad
 9/29/2021 12:14:16 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 29 03:44:32 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS092021.M
 Quant Title : 8080.M
 QLast Update : Tue Sep 21 01:52:13 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.158	6.747	513.5E6	193.1E6	337.056	329.261
Target Compounds							
1) T	Dalapon	2.996	2.502	573.3E6	263.0E6	240.451	361.779 #
2) T	3,5-DICHL...	6.382	5.944	544.2E6	204.6E6	255.702	232.028
3) T	4-Nitroph...	6.989	6.372	260.2E6	82925492	300.382	265.171m
5) T	DICAMBA	7.348	6.908	1605.5E6	672.4E6	275.080	263.970m
6) T	MCPD	7.512	7.000	96254449	39454951	22.216m	19.260
7) T	MCPA	7.659	7.210	152.6E6	84309112	24.667	27.687
8) T	DICHLORPROP	8.037	7.526	313.9E6	142.3E6	212.586	219.416
9) T	2,4-D	8.266	7.816	472.4E6	183.9E6	268.662	241.723
10) T	Pentachlo...	8.571	8.241	4093.4E6	1705.9E6	202.581	207.298
11) T	2,4,5-TP ...	9.120	8.601	1532.1E6	686.1E6	191.871	195.973
12) T	2,4,5-T	9.403	8.978	1659.3E6	669.1E6	201.037	205.619
13) T	2,4-DB	9.955	9.482	223.5E6	66443820	135.207	159.462
14) T	DINOSEB	11.103	9.808	362.2E6	167.3E6	59.445	68.310
15) T	Picloram	10.925	10.785	3884.5E6	1239.6E6	326.379m	288.969
16) T	DCPA	11.372	10.735	3079.2E6	1469.1E6	299.892	337.039m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS092821\
Data File : PS016917.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Sep 2021 17:49
Operator : DD\AJ
Sample : M3870-01MS
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
TP-CMS

Manual Integrations
APPROVED

mohammad
9/29/2021 12:14:16 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 29 03:44:32 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS092021.M
Quant Title : 8080.M
QLast Update : Tue Sep 21 01:52:13 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

