

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS092920\
 Data File : PS012564.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Sep 2020 22:24
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 30 01:44:15 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS092220.M
 Quant Title : 8080.M
 QLast Update : Wed Sep 23 04:57:25 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.214	6.911	812.5E6	383.4E6	829.513	833.378
Target Compounds							
1) T	Dalapon	3.035	2.625	769.6E6	406.9E6	677.217	680.288
2) T	3,5-DICHL...	6.429	6.064	1021.9E6	535.2E6	757.841	750.626
3) T	4-Nitroph...	7.046	6.520	379.3E6	215.5E6	779.930	771.353
5) T	DICAMBA	7.405	7.079	3002.2E6	1486.0E6	778.239	779.264
6) T	MCPD	7.575	7.176	233.6E6	118.0E6	80.072	77.988
7) T	MCPA	7.724	7.391	325.5E6	182.2E6	78.035	75.990
8) T	DICHLORPROP	8.096	7.704	751.1E6	399.0E6	776.154	772.670
9) T	2,4-D	8.326	8.000	890.7E6	498.2E6	784.423	795.485
10) T	Pentachlo...	8.635	8.427	10665.1E6	4753.0E6	771.728	779.251
11) T	2,4,5-TP ...	9.184	8.788	4351.5E6	2098.0E6	784.785	795.197
12) T	2,4,5-T	9.470	9.167	4406.1E6	2084.2E6	783.833	810.563
13) T	2,4-DB	10.020	9.673	838.7E6	294.3E6	886.284	786.096
14) T	DINOSEB	11.180	10.005	3398.0E6	1536.9E6	748.418	802.376
15) T	Picloram	11.004	10.991	5685.9E6	2996.5E6	789.690	760.247
16) T	DCPA	11.446	10.933	5538.5E6	2707.6E6	784.129	814.203

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS092920\
Data File : PS012564.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Sep 2020 22:24
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 30 01:44:15 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS092220.M
Quant Title : 8080.M
QLast Update : Wed Sep 23 04:57:25 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

