

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS093019\
 Data File : PS006947.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Sep 2019 12:44
 Operator : SM\AJ
 Sample : K5038-07MSD (Sig #1); K50338-07MSD (Sig #2)
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 TS-1MSD

Manual Integrations
 APPROVED

Ankita
 10/2/2019 8:48:32 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 01 10:05:59 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS092019.M
 Quant Title : 8080.M
 QLast Update : Fri Sep 20 11:14:00 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.479	5.209	660.8E6	325.0E6	723.154	706.730
Target Compounds							
1) T	Dalapon	2.049	1.755	307.6E6	230.2E6	287.925	431.784m#
2) T	3,5-DICHL...	4.848	4.513	145.7E6	79376723	133.951	138.080m
3) T	4-Nitroph...	5.333	4.896	110.3E6	57526711	289.253	237.979
5) T	DICAMBA	5.632	5.338	579.0E6	266.9E6	179.193	158.577
6) T	MCPD	5.782	5.419	31220965	20114383	14.675m	17.317
7) T	MCPA	5.915	5.595	61247599	43026894	19.687m	23.377m
8) T	DICHLORPROP	6.260	5.876	104.9E6	73913883	136.591m	178.502 #
9) T	2,4-D	6.476	6.146	200.5E6	119.1E6	244.611	233.399
10) T	Pentachlo...	6.737	6.514	1830.2E6	908.2E6	153.568	144.348m
11) T	2,4,5-TP ...	7.270	6.875	653.1E6	340.7E6	143.739	142.225
12) T	2,4,5-T	7.542	7.235	704.4E6	350.5E6	138.430	162.166m
13) T	2,4-DB	8.073	7.722	106.2E6	86619619	183.083m	304.456 #
14) T	DINOSEB	9.162	8.032	316.6E6	158.1E6	83.366	91.775
15) T	Picloram	8.995	8.970	1543.6E6	709.0E6	230.186	194.787
16) T	DCPA	9.435	8.927	1068.6E6	401.6E6	183.118	126.379m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS093019\
Data File : PS006947.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Sep 2019 12:44
Operator : SM\AJ
Sample : K5038-07MSD (Sig #1); K50338-07MSD (Sig #2)
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
TS-1MSD

Manual Integrations
APPROVED

Ankita
10/2/2019 8:48:32 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 01 10:05:59 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS092019.M
Quant Title : 8080.M
QLast Update : Fri Sep 20 11:14:00 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

