

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS100118\
 Data File : PS001859.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Oct 2018 23:37
 Operator : EI\SJ
 Sample : J5171-04
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 02 04:55:15 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS091718.M
 Quant Title : 8080.M
 QLast Update : Tue Sep 18 03:01:16 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.458	10.955	122.2E6	120.6E6	544.771	474.736
Target Compounds							
1) T	Dalapon	2.888	2.368	2734363	12342683	8.409	41.781 #
2) T	3,5-DICHL...	0.000	9.196	0	25976739	N.D.	82.300 #
6) T	MCPD	12.145	0.000	27572230	0	43.269	N.D. #
15) T	Picloram	0.000	16.836	0	7481354	N.D.	4.399 #
16) T	DCPA	0.000	16.888	0	7834881	N.D.	4.248 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS100118\
Data File : PS001859.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Oct 2018 23:37
Operator : EI\SJ
Sample : J5171-04
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 02 04:55:15 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS091718.M
Quant Title : 8080.M
QLast Update : Tue Sep 18 03:01:16 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 2.888 min
Delta R.T.: -0.061 min
Response: 2734363
Conc: 8.41 ng/ml

Instrument :
ECD_S
ClientSampleId :




#1 Dalapon

R.T.: 2.368 min
Delta R.T.: -0.041 min
Response: 12342683
Conc: 41.78 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 0.000 min
Exp R.T. : 9.902 min
Response: 0
Conc: N.D.



#2 3,5-DICHLOROBENZOIC ACID


R.T.: 9.196 min
Delta R.T.: 0.070 min
Response: 25976739
Conc: 82.30 ng/ml



#4 2,4-DCAA


R.T.: 11.458 min
Delta R.T.: -0.022 min
Response: 122200719
Conc: 544.77 ng/ml

Instrument :
ECD_S
ClientSampleId :




#4 2,4-DCAA

R.T.: 10.955 min
Delta R.T.: -0.021 min
Response: 120590863
Conc: 474.74 ng/ml




#6 MCPP

R.T.: 12.145 min
Delta R.T.: 0.020 min
Response: 27572230
Conc: 43.27 ug/ml



#6 MCPP


R.T.: 0.000 min
Exp R.T. : 11.483 min
Response: 0
Conc: N.D.



#15 Picloram


R.T.: 0.000 min
Exp R.T. : 16.988 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :




#15 Picloram

R.T.: 16.836 min
Delta R.T.: -0.014 min
Response: 7481354
Conc: 4.40 ng/ml



#16 DCPA

R.T.: 0.000 min
Exp R.T. : 17.565 min
Response: 0
Conc: N.D.



#16 DCPA

R.T.: 16.888 min
Delta R.T.: -0.018 min
Response: 7834881
Conc: 4.25 ng/ml