

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS100225\
 Data File : PS031938.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Oct 2025 19:38
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 03 02:11:09 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS090925.M
 Quant Title : 8080.M
 QLast Update : Tue Sep 09 13:56:38 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.280	7.739	2762.2E6	699.4E6	766.956	769.463
Target Compounds							
1) T	Dalapon	2.667	2.690	3359.3E6	1729.1E6	581.113	613.655
2) T	3,5-DICHL...	6.448	6.693	3810.4E6	1141.3E6	714.439	748.393
3) T	4-Nitroph...	7.092	7.294	816.7E6	771.6E6	740.338	611.019
5) T	DICAMBA	7.463	7.935	11209.6E6	4725.3E6	718.715	726.209
6) T	MCPD	7.641	8.031	663.0E6	141.4E6	74.164	75.084
7) T	MCPA	7.790	8.279	873.9E6	215.1E6	78.657	76.780
8) T	DICHLORPROP	8.170	8.654	2568.2E6	1046.4E6	721.053	719.135
9) T	2,4-D	8.409	8.996	2722.3E6	1182.6E6	794.676	718.838
10) T	Pentachlo...	8.702	9.503	42655.1E6	31930.0E6	803.390	804.475
11) T	2,4,5-TP ...	9.278	9.887	15608.8E6	11132.5E6	744.209	739.491
12) T	2,4,5-T	9.576	10.316	14167.7E6	9891.7E6	774.267	735.062
13) T	2,4-DB	10.150	10.885	1812.2E6	643.2E6	776.847	601.067
14) T	DINOSEB	11.344	11.259	11065.7E6	7646.2E6	767.246	786.587
15) T	Picloram	11.177	12.381	12589.0E6	16427.6E6	698.959	693.638
16) T	DCPA	11.643	12.299	20953.3E6	18629.2E6	777.492	772.866

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS100225\
Data File : PS031938.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Oct 2025 19:38
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 03 02:11:09 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS090925.M
Quant Title : 8080.M
QLast Update : Tue Sep 09 13:56:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

