

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS100323\
 Data File : PS024496.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Oct 2023 20:10
 Operator : AR\AJ
 Sample : 04651-01MS
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 SP1-COMPMS

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 10/04/2023
 Supervised By : Ankita Jodhani 10/04/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 04 03:18:01 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS100323.M
 Quant Title : 8080.M
 QLast Update : Tue Oct 03 14:54:54 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.359	6.783	1009.9E6	335.5E6	490.876	485.338m
Target Compounds							
1) T	Dalapon	2.478	2.543	549.4E6	257.2E6	222.387	229.820m
2) T	3,5-DICHL...	5.734	5.969	1311.9E6	432.5E6	418.235	398.033
5) T	DICAMBA	6.501	6.945	3633.2E6	1180.5E6	405.012	398.652
6) T	MCP	6.650	7.040	252.3E6	77055026	39.827m	39.102
7) T	MCPA	6.772	7.247	318.1E6	137.8E6	35.695	42.992
8) T	DICHLORPROP	7.088	7.560	964.3E6	311.1E6	410.194	413.740
9) T	2,4-D	7.285	7.843	1185.5E6	409.8E6	429.063	445.195
10) T	Pentachlo...	7.514	8.268	15626.0E6	4950.3E6	480.581	454.314
11) T	2,4,5-TP ...	8.035	8.627	6036.1E6	1735.4E6	435.661	409.178
12) T	2,4,5-T	8.290	8.996	5777.8E6	1755.5E6	442.227	425.735
13) T	2,4-DB	8.796	9.499	798.9E6	209.9E6	391.051	411.826
14) T	DINOSEB	9.848	9.827	4273.6E6	1258.2E6	418.059	405.724
15) T	Picloram	9.672	10.779	7028.3E6	2411.3E6	403.083	388.022
16) T	DCPA	10.113	10.750	7521.8E6	2062.4E6	441.870	395.290

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS100323\
Data File : PS024496.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Oct 2023 20:10
Operator : AR\AJ
Sample : 04651-01MS
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
SP1-COMPMS

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 10/04/2023
Supervised By : Ankita Jodhani 10/04/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 04 03:18:01 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS100323.M
Quant Title : 8080.M
QLast Update : Tue Oct 03 14:54:54 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

