

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS100421\
 Data File : PS016942.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Oct 2021 12:27
 Operator : DD\AJ
 Sample : M3951-02MS
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 WC-20210924MS

Manual Integrations
 APPROVED

mohammad
 10/5/2021 12:16:58 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 04 13:30:12 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS092021.M
 Quant Title : 8080.M
 QLast Update : Tue Sep 21 01:52:13 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.155	6.748	1289.3E6	259.9E6	846.222	443.319 #
Target Compounds							
1) T	Dalapon	2.998	2.504	863.6E6	386.8E6	362.162	532.120 #
2) T	3,5-DICHL...	6.382	5.943	1005.7E6	395.9E6	472.574	449.095m
5) T	DICAMBA	7.349	6.910	2580.2E6	1114.8E6	442.085	437.612
6) T	MCP	7.515	7.004	291.0E6	99561077	67.162m	48.601 #
7) T	MCPA	7.663	7.213	298.0E6	160.3E6	48.191	52.658
8) T	DICHLORPROP	8.037	7.526	685.2E6	298.4E6	464.060	460.096
9) T	2,4-D	8.265	7.816	869.7E6	370.8E6	494.620	487.316
10) T	Pentachlo...	8.572	8.241	9535.6E6	3847.1E6	471.909	467.504
11) T	2,4,5-TP ...	9.120	8.601	4098.9E6	2239.2E6	513.302	639.578
12) T	2,4,5-T	9.404	8.978	3975.8E6	1548.0E6	481.706	475.693
13) T	2,4-DB	9.953	9.481	558.3E6	196.1E6	337.765	470.579 #
14) T	DINOSEB	11.104	9.807	2499.3E6	1023.1E6	410.226m	417.661
15) T	Picloram	10.925	10.785	5396.5E6	1735.7E6	453.418m	404.619
16) T	DCPA	11.372	10.736	4657.1E6	2045.4E6	453.567	469.262m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS100421\
Data File : PS016942.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Oct 2021 12:27
Operator : DD\AJ
Sample : M3951-02MS
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
WC-20210924MS

Manual Integrations
APPROVED

mohammad
10/5/2021 12:16:58 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 04 13:30:12 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS092021.M
Quant Title : 8080.M
QLast Update : Tue Sep 21 01:52:13 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

