

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS100423\
Data File : PS024545.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Oct 2023 23:42
Operator : AR\AJ
Sample : PB156106BL
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
PB156106BL

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/05/2023
Supervised By :Ankita Jodhani 10/05/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 05 00:20:16 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS100323.M
Quant Title : 8080.M
QLast Update : Tue Oct 03 14:54:54 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
4) S 2,4-DCAA	6.357	6.782	1090.2E6	356.6E6	529.905m	515.886

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS100423\
Data File : PS024545.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Oct 2023 23:42
Operator : AR\AJ
Sample : PB156106BL
Misc :
ALS Vial : 14 Sample Multiplier: 1



Instrument :
ECD_S
ClientSampleId :
PB156106BL


Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 10/05/2023
Supervised By : Ankita Jodhani 10/05/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 05 00:20:16 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS100323.M
Quant Title : 8080.M
QLast Update : Tue Oct 03 14:54:54 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#4 2,4-DCAA

R.T.: 6.357 min
Delta R.T.: -0.002 min
Response: 1090188618
Conc: 529.90 ng/ml

Instrument :
ECD_S
ClientSampleId :
PB156106BL

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/05/2023
Supervised By :Ankita Jodhani 10/05/2023



#4 2,4-DCAA

R.T.: 6.782 min
Delta R.T.: -0.002 min
Response: 356627000
Conc: 515.89 ng/ml