

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS100520\
 Data File : PS012633.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Oct 2020 18:45
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 05 19:14:34 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS092220.M
 Quant Title : 8080.M
 QLast Update : Wed Sep 23 04:57:25 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.214	6.912	703.4E6	328.7E6	718.125	714.576
Target Compounds							
1) T	Dalapon	3.035	2.626	702.1E6	366.9E6	617.828	613.330
2) T	3,5-DICHL...	6.430	6.064	893.1E6	468.2E6	662.282	656.651
3) T	4-Nitroph...	7.048	6.521	311.8E6	179.9E6	641.272	644.025
5) T	DICAMBA	7.405	7.079	2633.0E6	1301.9E6	682.539	682.751
6) T	MCPD	7.572	7.174	203.9E6	103.2E6	69.895	68.238
7) T	MCPA	7.722	7.390	280.2E6	159.4E6	67.172	66.504
8) T	DICHLORPROP	8.096	7.704	650.6E6	347.0E6	672.272	672.079
9) T	2,4-D	8.327	8.001	756.6E6	422.3E6	666.340	674.380
10) T	Pentachlo...	8.635	8.427	9541.2E6	4286.7E6	690.398	702.802
11) T	2,4,5-TP ...	9.184	8.788	3807.4E6	1841.6E6	686.658	697.992
12) T	2,4,5-T	9.471	9.168	3813.4E6	1786.0E6	678.390	694.600
13) T	2,4-DB	10.024	9.675	679.8E6	242.9E6	718.340	648.767
14) T	DINOSEB	11.179	10.005	3003.1E6	1311.1E6	661.445	684.491
15) T	Picloram	11.010	10.996	4399.7E6	2456.3E6	611.053	623.192
16) T	DCPA	11.445	10.932	4862.0E6	2376.8E6	688.351	714.737

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS100520\
Data File : PS012633.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Oct 2020 18:45
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 05 19:14:34 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS092220.M
Quant Title : 8080.M
QLast Update : Wed Sep 23 04:57:25 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

