

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS100622\
 Data File : PS020743.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Oct 2022 21:43
 Operator : AR\AJ
 Sample : HSTDICC1000
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 07 06:06:48 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS100622.M
 Quant Title : 8080.M
 QLast Update : Fri Oct 07 05:57:06 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.422	6.868	1824.2E6	1319.4E6	959.347	986.033
Target Compounds							
1) T	Dalapon	2.510	2.585	2551.7E6	1588.8E6	916.418	924.120
2) T	3,5-DICHL...	5.787	6.035	2668.1E6	1835.0E6	874.641	911.767
3) T	4-Nitroph...	6.256	6.473	1205.9E6	745.8E6	912.612	906.925
5) T	DICAMBA	6.569	7.036	7245.9E6	5457.9E6	900.131	920.556
6) T	MCPD	6.732	7.140	563.6E6	420.9E6	97.439	97.755
7) T	MCPA	6.860	7.354	792.1E6	577.4E6	93.640	96.892
8) T	DICHLORPROP	7.171	7.665	1828.4E6	1348.6E6	905.023	931.251
9) T	2,4-D	7.369	7.955	2233.1E6	1651.5E6	910.062	930.816
10) T	Pentachlo...	7.607	8.393	24389.7E6	18000.6E6	855.317	862.915
11) T	2,4,5-TP ...	8.135	8.755	10294.7E6	7813.8E6	896.954	908.907
12) T	2,4,5-T	8.392	9.131	10157.4E6	7539.6E6	898.129	914.808
13) T	2,4-DB	8.903	9.640	2133.8E6	1062.5E6	932.429	957.379
14) T	DINOSEB	9.967	9.973	7467.3E6	5611.0E6	902.430	922.780
15) T	Picloram	9.789	10.944	15410.5E6	11049.8E6	912.112	926.345
16) T	DCPA	10.234	10.907	12689.4E6	9463.4E6	891.938	903.328

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS100622\
Data File : PS020743.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Oct 2022 21:43
Operator : AR\AJ
Sample : HSTDICC1000
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 07 06:06:48 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS100622.M
Quant Title : 8080.M
QLast Update : Fri Oct 07 05:57:06 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

