

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS100720\
 Data File : PS012711.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Oct 2020 14:24
 Operator : DD\AJ
 Sample : PB132126BL
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 07 16:53:13 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS092220.M
 Quant Title : 8080.M
 QLast Update : Wed Sep 23 04:57:25 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.217	6.913	514.9E6	274.0E6	525.656	595.649
Target Compounds							
1) T	Dalapon	2.993	2.588	63803357	110.1E6	56.147	184.132 #
2) T	3,5-DICHL...	0.000	6.095	0	72151910	N.D.	101.200
3) T	4-Nitroph...	6.991	0.000	18404180	0	37.848	N.D. #
5) T	DICAMBA	7.394	7.047	8711185	9099416	2.258	4.772 #
6) T	MCPD	7.569	0.000	11047484	0	3.786	N.D. #
7) T	MCPA	7.762	7.372	63850381	46405778	15.309	19.358 #
8) T	DICHLORPROP	0.000	7.747	0	7939566	N.D.	15.377
9) T	2,4-D	8.397	7.973	45884353	6373459	40.412	10.177 #
10) T	Pentachlo...	8.692	0.000	20713211	0	1.499	N.D. #
11) T	2,4,5-TP ...	0.000	8.763	0	38936748	N.D.	14.758
12) T	2,4,5-T	9.408	0.000	21449806	0	3.816	N.D. #
14) T	DINOSEB	0.000	10.007	0	7991987	N.D.	4.172
15) T	Picloram	10.991	11.008	18705998	14221329	2.598	3.608 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS100720\
Data File : PS012711.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Oct 2020 14:24
Operator : DD\AJ
Sample : PB132126BL
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 07 16:53:13 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS092220.M
Quant Title : 8080.M
QLast Update : Wed Sep 23 04:57:25 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 2.993 min
Delta R.T.: -0.039 min
Response: 63803357
Conc: 56.15 ng/ml

Instrument :
ECD_S
ClientSampleId :




#1 Dalapon

R.T.: 2.588 min
Delta R.T.: -0.032 min
Response: 110146137
Conc: 184.13 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 0.000 min
Exp R.T. : 6.430 min
Response: 0
Conc: N.D.



#2 3,5-DICHLOROBENZOIC ACID


R.T.: 6.095 min
Delta R.T.: 0.030 min
Response: 72151910
Conc: 101.20 ng/ml



#3 4-Nitrophenol


R.T.: 6.991 min
Delta R.T.: -0.056 min
Response: 18404180
Conc: 37.85 ng/ml

Instrument :
ECD_S
ClientSampleId :




#3 4-Nitrophenol

R.T.: 0.000 min
Exp R.T. : 6.522 min
Response: 0
Conc: N.D.




#4 2,4-DCAA

R.T.: 7.217 min
Delta R.T.: 0.003 min
Response: 514852944
Conc: 525.66 ng/ml



#4 2,4-DCAA


R.T.: 6.913 min
Delta R.T.: 0.000 min
Response: 274008233
Conc: 595.65 ng/ml



#5 DICAMBA


R.T.: 7.394 min
Delta R.T.: -0.011 min
Response: 8711185
Conc: 2.26 ng/ml

Instrument :
ECD_S
ClientSampleId :




#5 DICAMBA

R.T.: 7.047 min
Delta R.T.: -0.032 min
Response: 9099416
Conc: 4.77 ng/ml




#6 MCP

R.T.: 7.569 min
Delta R.T.: -0.003 min
Response: 11047484
Conc: 3.79 ug/ml



#6 MCP


R.T.: 0.000 min
Exp R.T.: 7.175 min
Response: 0
Conc: N.D.



#7 MCPA


R.T.: 7.762 min
Delta R.T.: 0.040 min
Response: 63850381
Conc: 15.31 ug/ml

Instrument :
ECD_S
ClientSampleId :




#7 MCPA

R.T.: 7.372 min
Delta R.T.: -0.017 min
Response: 46405778
Conc: 19.36 ug/ml




#8 DICHLORPROP

R.T.: 0.000 min
Exp R.T. : 8.097 min
Response: 0
Conc: N.D.



#8 DICHLORPROP


R.T.: 7.747 min
Delta R.T.: 0.042 min
Response: 7939566
Conc: 15.38 ng/ml



#9 2,4-D


R.T.: 8.397 min
Delta R.T.: 0.069 min
Response: 45884353
Conc: 40.41 ng/ml

Instrument :
ECD_S
ClientSampleId :




#9 2,4-D

R.T.: 7.973 min
Delta R.T.: -0.028 min
Response: 6373459
Conc: 10.18 ng/ml




#10 Pentachlorophenol

R.T.: 8.692 min
Delta R.T.: 0.056 min
Response: 20713211
Conc: 1.50 ng/ml



#10 Pentachlorophenol


R.T.: 0.000 min
Exp R.T. : 8.428 min
Response: 0
Conc: N.D.



#11 2,4,5-TP (SILVEX)


R.T.: 0.000 min
Exp R.T.: 9.185 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :




#11 2,4,5-TP (SILVEX)

R.T.: 8.763 min
Delta R.T.: -0.026 min
Response: 38936748
Conc: 14.76 ng/ml




#12 2,4,5-T

R.T.: 9.408 min
Delta R.T.: -0.063 min
Response: 21449806
Conc: 3.82 ng/ml



#12 2,4,5-T


R.T.: 0.000 min
Exp R.T.: 9.169 min
Response: 0
Conc: N.D.



#14 DINOSEB


R.T.: 0.000 min
Exp R.T.: 11.181 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :




#14 DINOSEB

R.T.: 10.007 min
Delta R.T.: 0.000 min
Response: 7991987
Conc: 4.17 ng/ml



#15 Picloram

R.T.: 10.991 min
Delta R.T.: -0.018 min
Response: 18705998
Conc: 2.60 ng/ml



#15 Picloram

R.T.: 11.008 min
Delta R.T.: 0.013 min
Response: 14221329
Conc: 3.61 ng/ml