

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS100720\
 Data File : PS012749.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Oct 2020 04:58
 Operator : DD\AJ
 Sample : L4260-04MSD
 Misc :
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 PAV-B21-100220-0-10MSD

Manual Integrations
 APPROVED

Ankita
 10/8/2020 11:15:58 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 08 06:33:30 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS092220.M
 Quant Title : 8080.M
 QLast Update : Wed Sep 23 04:57:25 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.214	6.911	338.0E6	182.8E6	345.140	397.298
Target Compounds							
1) T	Dalapon	3.033	2.625	491.3E6	396.1E6	432.378	662.203 #
2) T	3,5-DICHL...	6.429	6.064	643.3E6	308.7E6	477.057	432.995
5) T	DICAMBA	7.404	7.078	1816.8E6	900.0E6	470.956	471.947
6) T	MCP	7.570	7.172	151.3E6	70358052	51.842	46.513
7) T	MCPA	7.719	7.385	304.1E6	127.0E6	72.920m	52.974m#
8) T	DICHLORPROP	8.095	7.703	477.1E6	254.7E6	493.072	493.224
9) T	2,4-D	8.326	7.999	587.2E6	337.7E6	517.125	539.211
10) T	Pentachlo...	8.634	8.426	5308.4E6	2447.6E6	384.116	401.288
11) T	2,4,5-TP ...	9.183	8.784	2919.0E6	2106.3E6	526.432	798.342 #
12) T	2,4,5-T	9.469	9.167	2845.7E6	1348.2E6	506.246	524.319
13) T	2,4-DB	10.023	9.673	303.9E6	181.6E6	321.198	485.098 #
14) T	DINOSEB	11.179	10.004	1322.4E6	555.6E6	291.268	290.040
15) T	Picloram	11.007	10.993	2972.3E6	1630.5E6	412.810	413.676
16) T	DCPA	11.445	10.932	3116.9E6	1556.4E6	441.280	468.033

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS100720\
Data File : PS012749.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Oct 2020 04:58
Operator : DD\AJ
Sample : L4260-04MSD
Misc :
ALS Vial : 38 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
PAV-B21-100220-0-10MSD

Manual Integrations
APPROVED

Ankita
10/8/2020 11:15:58 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 08 06:33:30 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS092220.M
Quant Title : 8080.M
QLast Update : Wed Sep 23 04:57:25 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

