

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS100721\
 Data File : PS017020.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Oct 2021 14:21
 Operator : DD\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 08 01:30:47 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS092021.M
 Quant Title : 8080.M
 QLast Update : Tue Sep 21 01:52:13 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.160	6.747	659.7E6	285.3E6	433.033	486.548
Target Compounds							
2) T	3,5-DICHL...	6.423	0.000	3340924	0	1.570	N.D. #
7) T	MCPA	7.631f	0.000	12226641	0	1.977	N.D. #
9) T	2,4-D	8.340f	0.000	101.0E6	0	57.426	N.D. #
15) T	Picloram	10.907	0.000	7034877	0	<MDL	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS100721\
Data File : PS017020.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Oct 2021 14:21
Operator : DD\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 08 01:30:47 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS092021.M
Quant Title : 8080.M
QLast Update : Tue Sep 21 01:52:13 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#2 3,5-DICHLOROBENZOIC ACID


R.T.: 6.423 min
Delta R.T.: 0.032 min
Response: 3340924
Conc: 1.57 ng/ml

Instrument :
ECD_S
ClientSampleId :




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 0.000 min
Exp R.T. : 5.951 min
Response: 0
Conc: N.D.




#4 2,4-DCAA

R.T.: 7.160 min
Delta R.T.: -0.010 min
Response: 659747017
Conc: 433.03 ng/ml



#4 2,4-DCAA


R.T.: 6.747 min
Delta R.T.: -0.012 min
Response: 285293843
Conc: 486.55 ng/ml



#7 MCPA


R.T.: 7.631 min
Delta R.T.: -0.047 min
Response: 12226641
Conc: 1.98 ug/ml

Instrument :
ECD_S
ClientSampleId :




#7 MCPA

R.T.: 0.000 min
Exp R.T. : 7.229 min
Response: 0
Conc: N.D.



#9 2,4-D

R.T.: 8.340 min
Delta R.T.: 0.062 min
Response: 100969635
Conc: 57.43 ng/ml



#9 2,4-D

R.T.: 0.000 min
Exp R.T. : 7.829 min
Response: 0
Conc: N.D.