

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS100721\
 Data File : PS017021.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Oct 2021 14:45
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 08 01:31:04 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS092021.M
 Quant Title : 8080.M
 QLast Update : Tue Sep 21 01:52:13 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.159	6.747	1148.5E6	455.0E6	753.824	775.897
Target Compounds							
1) T	Dalapon	2.999	2.502	1601.0E6	502.8E6	671.433	691.827
2) T	3,5-DICHL...	6.382	5.942	1476.7E6	619.9E6	693.895	703.132
3) T	4-Nitroph...	6.989	6.370	599.0E6	207.2E6	691.529	662.553
5) T	DICAMBA	7.349	6.909	4242.3E6	1887.1E6	726.865	740.777
6) T	MCPD	7.518	7.006	281.3E6	151.3E6	64.923	73.851
7) T	MCPA	7.667	7.216	453.6E6	215.0E6	73.345	70.589
8) T	DICHLORPROP	8.038	7.524	1057.0E6	462.3E6	715.895	712.782
9) T	2,4-D	8.266	7.814	1263.8E6	519.1E6	718.782	682.302
10) T	Pentachlo...	8.572	8.241	14843.6E6	6105.3E6	734.600	741.918
11) T	2,4,5-TP ...	9.120	8.600	5824.3E6	2599.2E6	729.384	742.387
12) T	2,4,5-T	9.403	8.976	6062.0E6	2381.0E6	734.476	731.651
13) T	2,4-DB	9.952	9.480	1160.4E6	310.2E6	701.962	744.416
14) T	DINOSEB	11.104	9.805	4404.8E6	1854.6E6	722.978	757.136
15) T	Picloram	10.925	10.782	8420.3E6	3190.8E6	707.480	743.818
16) T	DCPA	11.371	10.734	7407.3E6	3236.0E6	721.411	742.409

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS100721\
Data File : PS017021.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Oct 2021 14:45
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 08 01:31:04 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS092021.M
Quant Title : 8080.M
QLast Update : Tue Sep 21 01:52:13 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

