

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS100919\
 Data File : PS007110.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Oct 2019 13:16
 Operator : SM\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Ankita
 10/9/2019 4:08:09 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 09 14:02:03 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS092019.M
 Quant Title : 8080.M
 QLast Update : Fri Sep 20 11:14:00 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.461	5.191	713.0E6	346.4E6	780.248	753.255
Target Compounds							
1) T	Dalapon	2.032	1.740	761.5E6	378.0E6	712.729m	709.190m
2) T	3,5-DICHL...	4.828	4.493	762.5E6	375.7E6	701.038m	653.556m
3) T	4-Nitroph...	5.318	4.878	275.6E6	168.6E6	722.883	697.656
5) T	DICAMBA	5.612	5.317	2432.0E6	1167.4E6	752.698	693.547m
6) T	MCPD	5.763	5.401	167.0E6	84688866	78.494	72.912
7) T	MCPA	5.896	5.578	239.3E6	129.8E6	76.911	70.520
8) T	DICHLORPROP	6.241	5.856	584.4E6	291.7E6	761.099	704.543
9) T	2,4-D	6.458	6.129	637.5E6	358.7E6	777.937m	703.163
10) T	Pentachlo...	6.716	6.493	9315.5E6	4567.1E6	781.631	725.917
11) T	2,4,5-TP ...	7.251	6.856	3544.6E6	1741.3E6	780.177	726.960
12) T	2,4,5-T	7.525	7.215	3974.5E6	1513.2E6	781.113	700.088m
13) T	2,4-DB	8.055	7.702	439.1E6	205.8E6	757.099m	723.462
14) T	DINOSB	9.141	8.011	3032.7E6	1257.5E6	798.506	729.960
15) T	Picloram	8.982	8.951	4510.0E6	2286.0E6	672.569	628.073
16) T	DCPA	9.413	8.906	4643.0E6	2330.4E6	795.652	733.423m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS100919\
Data File : PS007110.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Oct 2019 13:16
Operator : SM\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Ankita
10/9/2019 4:08:09 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 09 14:02:03 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS092019.M
Quant Title : 8080.M
QLast Update : Fri Sep 20 11:14:00 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

