

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS100920\
 Data File : PS012784.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Oct 2020 19:47
 Operator : DD\AJ
 Sample : L4292-04
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 PAV-B13-100620-0-10

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 10 08:48:25 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS092220.M
 Quant Title : 8080.M
 QLast Update : Wed Sep 23 04:57:25 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.215	6.912	565.0E6	292.4E6	576.869	635.712
Target Compounds							
1) T	Dalapon	2.993	2.590	61633814	156.4E6	54.238	261.457 #
2) T	3,5-DICHL...	6.473	6.094	9933990	73443959	7.367	103.013 #
3) T	4-Nitroph...	7.052	6.574	15690260	-2522776	32.267	N.D. #
5) T	DICAMBA	7.392	7.045	10827321	18841157	2.807	9.880 #
6) T	MCPD	7.569	0.000	8017609	0	2.748	N.D. #
7) T	MCPA	7.717	7.350	55608560	10341411	13.333	4.314 #
8) T	DICHLORPROP	0.000	7.741	0	18696796	N.D.	36.211
10) T	Pentachlo...	8.689	0.000	8409699	0	<MDL	N.D. #
11) T	2,4,5-TP ...	0.000	8.760	0	14574877	N.D.	5.524
12) T	2,4,5-T	9.405	0.000	24330014	0	4.328	N.D. #
14) T	DINOSEB	11.152	10.050	11465476	56238290	2.525	29.361 #
15) T	Picloram	10.980	0.000	61918154	0	8.600	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS100920\
Data File : PS012784.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Oct 2020 19:47
Operator : DD\AJ
Sample : L4292-04
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
PAV-B13-100620-0-10

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 08:48:25 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS092220.M
Quant Title : 8080.M
QLast Update : Wed Sep 23 04:57:25 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 2.993 min
Delta R.T.: -0.039 min
Response: 61633814
Conc: 54.24 ng/ml

Instrument :
ECD_S
ClientSampleId :
PAV-B13-100620-0-10




#1 Dalapon

R.T.: 2.590 min
Delta R.T.: -0.031 min
Response: 156400880
Conc: 261.46 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 6.473 min
Delta R.T.: 0.043 min
Response: 9933990
Conc: 7.37 ng/ml



#2 3,5-DICHLOROBENZOIC ACID


R.T.: 6.094 min
Delta R.T.: 0.029 min
Response: 73443959
Conc: 103.01 ng/ml



#3 4-Nitrophenol


R.T.: 7.052 min
Delta R.T.: 0.005 min
Response: 15690260
Conc: 32.27 ng/ml

Instrument :
ECD_S
ClientSampleId :
PAV-B13-100620-0-10




#3 4-Nitrophenol

R.T.: 6.574 min
Delta R.T.: 0.052 min
Response: -2522776
Conc: N.D.




#4 2,4-DCAA

R.T.: 7.215 min
Delta R.T.: 0.000 min
Response: 565013687
Conc: 576.87 ng/ml



#4 2,4-DCAA


R.T.: 6.912 min
Delta R.T.: 0.000 min
Response: 292438006
Conc: 635.71 ng/ml



#5 DICAMBA


R.T.: 7.392 min
Delta R.T.: -0.013 min
Response: 10827321
Conc: 2.81 ng/ml

Instrument :
ECD_S
ClientSampleId :
PAV-B13-100620-0-10




#5 DICAMBA

R.T.: 7.045 min
Delta R.T.: -0.034 min
Response: 18841157
Conc: 9.88 ng/ml




#6 MCP

R.T.: 7.569 min
Delta R.T.: -0.004 min
Response: 8017609
Conc: 2.75 ug/ml



#6 MCP


R.T.: 0.000 min
Exp R.T.: 7.175 min
Response: 0
Conc: N.D.



#7 MCPA


R.T.: 7.717 min
Delta R.T.: -0.005 min
Response: 55608560
Conc: 13.33 ug/ml

Instrument :
ECD_S
ClientSampleId :
PAV-B13-100620-0-10




#7 MCPA

R.T.: 7.350 min
Delta R.T.: -0.040 min
Response: 10341411
Conc: 4.31 ug/ml




#8 DICHLORPROP

R.T.: 0.000 min
Exp R.T. : 8.097 min
Response: 0
Conc: N.D.



#8 DICHLORPROP


R.T.: 7.741 min
Delta R.T.: 0.036 min
Response: 18696796
Conc: 36.21 ng/ml



#11 2,4,5-TP (SILVEX)


R.T.: 0.000 min
Exp R.T.: 9.185 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :
PAV-B13-100620-0-10




#11 2,4,5-TP (SILVEX)

R.T.: 8.760 min
Delta R.T.: -0.029 min
Response: 14574877
Conc: 5.52 ng/ml




#12 2,4,5-T

R.T.: 9.405 min
Delta R.T.: -0.066 min
Response: 24330014
Conc: 4.33 ng/ml



#12 2,4,5-T


R.T.: 0.000 min
Exp R.T.: 9.169 min
Response: 0
Conc: N.D.



#14 DINOSEB


R.T.: 11.152 min
Delta R.T.: -0.029 min
Response: 11465476
Conc: 2.53 ng/ml

Instrument :
ECD_S
ClientSampleId :
PAV-B13-100620-0-10




#14 DINOSEB

R.T.: 10.050 min
Delta R.T.: 0.044 min
Response: 56238290
Conc: 29.36 ng/ml



#15 Picloram

R.T.: 10.980 min
Delta R.T.: -0.029 min
Response: 61918154
Conc: 8.60 ng/ml



#15 Picloram

R.T.: 0.000 min
Exp R.T. : 10.995 min
Response: 0
Conc: N.D.