

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS100920\
 Data File : PS012787.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Oct 2020 20:56
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 10 08:49:17 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS092220.M
 Quant Title : 8080.M
 QLast Update : Wed Sep 23 04:57:25 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.214	6.911	724.3E6	329.3E6	739.533	715.932
Target Compounds							
1) T	Dalapon	3.035	2.625	669.3E6	343.3E6	588.959	573.816
2) T	3,5-DICHL...	6.430	6.065	925.8E6	475.4E6	686.536	666.819
3) T	4-Nitroph...	7.047	6.522	320.3E6	180.2E6	658.781	644.951
5) T	DICAMBA	7.404	7.078	2708.7E6	1312.5E6	702.148	688.288
6) T	MCPD	7.572	7.173	209.2E6	103.8E6	71.705	68.618
7) T	MCPA	7.721	7.389	287.1E6	159.0E6	68.826	66.333
8) T	DICHLORPROP	8.095	7.703	671.9E6	348.8E6	694.337	675.551
9) T	2,4-D	8.326	8.000	778.1E6	421.4E6	685.255	672.945
10) T	Pentachlo...	8.634	8.426	9874.0E6	4318.9E6	714.480	708.084
11) T	2,4,5-TP ...	9.183	8.787	3914.5E6	1846.5E6	705.976	699.859
12) T	2,4,5-T	9.470	9.167	3930.2E6	1787.5E6	699.175	695.161
13) T	2,4-DB	10.023	9.674	672.5E6	242.5E6	710.671	647.820
14) T	DINOSEB	11.178	10.004	3168.0E6	1324.1E6	697.759	691.266
15) T	Picloram	11.009	10.994	4632.6E6	2548.2E6	643.407	646.525
16) T	DCPA	11.444	10.931	5085.3E6	2402.9E6	719.971	722.599

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS100920\
Data File : PS012787.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Oct 2020 20:56
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 08:49:17 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS092220.M
Quant Title : 8080.M
QLast Update : Wed Sep 23 04:57:25 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

