

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS100923\
 Data File : PS024595.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Oct 2023 16:23
 Operator : AR\AJ
 Sample : PB156211BS
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 PB156211BS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 09 23:53:18 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS100323.M
 Quant Title : 8080.M
 QLast Update : Tue Oct 03 14:54:54 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.359	6.781	1144.6E6	375.0E6	556.365	542.500
Target Compounds							
1) T	Dalapon	2.478	2.539	1224.9E6	526.6E6	495.809	470.574
2) T	3,5-DICHL...	5.733	5.967	1607.4E6	552.3E6	512.424	508.277
3) T	4-Nitroph...	6.196	6.395	679.0E6	227.3E6	501.561	498.362
5) T	DICAMBA	6.500	6.943	4742.2E6	1523.0E6	528.639	514.292
6) T	MCPD	6.651	7.038	317.2E6	97590473	50.062	49.523
7) T	MCPA	6.773	7.245	433.7E6	158.9E6	48.660	49.574
8) T	DICHLORPROP	7.089	7.557	1232.1E6	392.8E6	524.111	522.327
9) T	2,4-D	7.286	7.841	1455.2E6	472.1E6	526.673	512.908
10) T	Pentachlo...	7.515	8.265	18038.8E6	5792.7E6	554.787	531.629
11) T	2,4,5-TP ...	8.037	8.623	7498.7E6	2216.1E6	541.221	522.529
12) T	2,4,5-T	8.292	8.994	7031.0E6	2147.8E6	538.143	520.872
13) T	2,4-DB	8.797	9.497	1085.3E6	271.0E6	531.238	531.785
14) T	DINOSEB	9.851	9.825	5361.4E6	1583.8E6	524.469	510.743
15) T	Picloram	9.677	10.777	8737.5E6	2995.2E6	501.112	481.978
16) T	DCPA	10.116	10.747	9299.6E6	2803.8E6	546.302	537.387

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS100923\
Data File : PS024595.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Oct 2023 16:23
Operator : AR\AJ
Sample : PB156211BS
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
PB156211BS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 09 23:53:18 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS100323.M
Quant Title : 8080.M
QLast Update : Tue Oct 03 14:54:54 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

