

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS101022\
 Data File : PS020765.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Oct 2022 10:41
 Operator : AR\AJ
 Sample : N4955-09
 Misc : MDL 100PPB
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 MDL-WATER-03-QT4-2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 11 00:06:24 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS100622.M
 Quant Title : 8080.M
 QLast Update : Fri Oct 07 06:09:52 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.422	6.863	1010.7E6	681.7E6	535.774	509.145
Target Compounds							
1) T	Dalapon	2.510	2.577	354.4E6	285.6E6	126.358	164.536 #
2) T	3,5-DICHL...	5.787	6.030	303.2E6	179.1E6	100.930	89.339
3) T	4-Nitroph...	6.256	6.468	154.8E6	70581421	112.031	85.602
5) T	DICAMBA	6.568	7.031	780.4E6	531.6E6	98.160	90.320
6) T	MCPD	6.722	7.125	78821905	35845117	13.413	8.180 #
7) T	MCPA	6.848	7.335	81082627	66838464	9.519	11.038
8) T	DICHLORPROP	7.171	7.659	218.1E6	159.9E6	108.724	110.259
9) T	2,4-D	7.370	7.949	271.0E6	259.7E6	111.135	146.520 #
10) T	Pentachlo...	7.606	8.387	3273.8E6	2195.5E6	119.230	108.688
11) T	2,4,5-TP ...	8.137	8.751	1361.6E6	1010.8E6	120.772	119.303
12) T	2,4,5-T	8.399	9.127	1360.4E6	779.7E6	122.394	95.830
13) T	2,4-DB	8.906	9.637	183.4E6	120.4E6	80.313	107.492 #
14) T	DINOSEB	9.969	9.968	826.8E6	572.7E6	100.968	94.503
15) T	Picloram	9.790	10.938	1862.9E6	1098.2E6	111.862	93.161
16) T	DCPA	10.235	10.902	1645.6E6	1058.5E6	118.292	103.027

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS101022\
Data File : PS020765.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2022 10:41
Operator : AR\AJ
Sample : N4955-09
Misc : MDL 100PPB
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
MDL-WATER-03-QT4-2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 11 00:06:24 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS100622.M
Quant Title : 8080.M
QLast Update : Fri Oct 07 06:09:52 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

