

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS101024\
 Data File : PS027899.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Oct 2024 10:09
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 10 11:45:46 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS100724.M
 Quant Title : 8080.M
 QLast Update : Mon Oct 07 13:46:03 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.096	7.625	1970.8E6	718.0E6	731.466	722.573
Target Compounds							
1) T	Dalapon	2.545	2.618	2996.5E6	1073.5E6	667.684	644.480
2) T	3,5-DICHL...	6.287	6.595	2687.2E6	937.5E6	676.507	685.574
3) T	4-Nitroph...	6.890	7.159	1226.4E6	435.0E6	671.690	640.838
5) T	DICAMBA	7.274	7.819	7650.3E6	2431.1E6	701.276	699.172
6) T	MCPD	7.454	7.922	534.9E6	191.1E6	71.924	69.444
7) T	MCPA	7.599	8.162	741.7E6	325.2E6	70.027	65.011
8) T	DICHLORPROP	7.960	8.525	1980.4E6	661.7E6	692.007	670.473
9) T	2,4-D	8.182	8.849	2300.0E6	699.3E6	683.701	697.723
10) T	Pentachlo...	8.468	9.364	27601.8E6	9230.8E6	710.402	667.977
11) T	2,4,5-TP ...	9.034	9.739	11237.4E6	3381.4E6	712.775	682.533
12) T	2,4,5-T	9.319	10.153	11544.9E6	2823.1E6	713.854	655.728
13) T	2,4-DB	9.882	10.716	1823.0E6	327.3E6	702.870	646.396
14) T	DINOSEB	11.058	11.089	7504.8E6	2391.4E6	711.831	707.918
15) T	Picloram	10.873	12.165	15157.9E6	3011.7E6	704.120	720.940
16) T	DCPA	11.357	12.123	12784.2E6	3369.3E6	719.794	701.885

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS101024\
Data File : PS027899.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2024 10:09
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 11:45:46 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS100724.M
Quant Title : 8080.M
QLast Update : Mon Oct 07 13:46:03 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

