

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS101325\
 Data File : PS031990.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Oct 2025 13:41
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 13 14:02:22 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS101325.M
 Quant Title : 8080.M
 QLast Update : Mon Oct 13 12:51:14 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.282	7.744	3588.6E6	867.9E6	752.793	788.776
Target Compounds							
1) T	Dalapon	2.664	2.688	5199.6E6	2476.0E6	672.862	658.808
2) T	3,5-DICHL...	6.449	6.696	4931.8E6	1460.8E6	700.647	700.319
3) T	4-Nitroph...	7.099	7.299	997.7E6	994.7E6	747.513	655.551
5) T	DICAMBA	7.465	7.939	14890.3E6	6259.3E6	711.014	722.704
6) T	MCPD	7.643	8.036	911.5E6	185.9E6	74.536	74.892
7) T	MCPA	7.792	8.284	1093.1E6	282.1E6	74.281	74.774
8) T	DICHLORPROP	8.173	8.658	3321.2E6	1394.5E6	698.592	712.115
9) T	2,4-D	8.412	9.002	3495.7E6	1491.3E6	719.900	717.576
10) T	Pentachlo...	8.710	9.507	48582.0E6	40032.4E6	723.814	772.470
11) T	2,4,5-TP ...	9.280	9.893	20367.1E6	14633.6E6	721.164	728.953
12) T	2,4,5-T	9.578	10.322	18688.2E6	12842.0E6	740.293	749.037
13) T	2,4-DB	10.154	10.890	2358.0E6	934.0E6	760.564	722.870
14) T	DINOSEB	11.347	11.264	15622.4E6	10998.9E6	700.531	723.463
15) T	Picloram	11.180	12.385	18909.6E6	25224.9E6	802.007	757.685
16) T	DCPA	11.646	12.304	26550.5E6	24243.0E6	722.655	729.954

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS101325\
Data File : PS031990.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Oct 2025 13:41
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 13 14:02:22 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS101325.M
Quant Title : 8080.M
QLast Update : Mon Oct 13 12:51:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

