

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS101419\
 Data File : PS007127.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Oct 2019 12:58
 Operator : SM\AJ
 Sample : HSTDICC1000
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 14 13:51:48 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS101419.M
 Quant Title : 8080.M
 QLast Update : Mon Oct 14 13:50:27 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.436	5.169	1337.4E6	693.0E6	990.513	997.780
Target Compounds							
1) T	Dalapon	2.011	1.721	1549.4E6	843.8E6	909.002	922.772
2) T	3,5-DICHL...	4.810	4.475	1576.5E6	805.1E6	918.447	930.192
3) T	4-Nitroph...	5.288	4.856	566.6E6	356.2E6	928.589	904.110
5) T	DICAMBA	5.586	5.295	4565.0E6	2429.0E6	927.618	928.442
6) T	MCPD	5.739	5.382	349.2E6	183.9E6	97.130	95.859
7) T	MCPA	5.871	5.558	485.2E6	270.6E6	95.478	94.366
8) T	DICHLORPROP	6.210	5.827	1099.1E6	608.1E6	936.392	936.764
9) T	2,4-D	6.423	6.096	1185.2E6	747.3E6	945.167	940.621
10) T	Pentachlo...	6.680	6.458	16279.5E6	8715.9E6	908.964	905.132
11) T	2,4,5-TP ...	7.214	6.822	6510.8E6	3556.7E6	931.699	933.205
12) T	2,4,5-T	7.486	7.179	7109.9E6	3358.8E6	925.749	943.155
13) T	2,4-DB	8.013	7.663	969.0E6	462.5E6	994.298	964.795
14) T	DINOSB	9.096	7.971	5320.4E6	2626.8E6	912.525	930.989
15) T	Picloram	8.927	8.905	10660.8E6	5835.0E6	941.184	939.824
16) T	DCPA	9.369	8.865	8343.9E6	4696.8E6	929.436	935.509

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS101419\
Data File : PS007127.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Oct 2019 12:58
Operator : SM\AJ
Sample : HSTDICC1000
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 14 13:51:48 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS101419.M
Quant Title : 8080.M
QLast Update : Mon Oct 14 13:50:27 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

