

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS101422\
 Data File : PS020798.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Oct 2022 09:33
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 15 00:14:07 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS100622.M
 Quant Title : 8080.M
 QLast Update : Fri Oct 07 06:09:52 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.419	6.861	1316.9E6	949.7E6	698.085	709.234
Target Compounds							
1) T	Dalapon	2.509	2.582	1833.6E6	1135.4E6	653.680	654.117
2) T	3,5-DICHL...	5.785	6.030	1961.6E6	1317.2E6	652.890	657.092
3) T	4-Nitroph...	6.253	6.467	879.1E6	531.0E6	656.674	643.983
5) T	DICAMBA	6.565	7.030	5344.3E6	4039.5E6	672.214	686.318
6) T	MCPD	6.725	7.130	394.5E6	301.7E6	67.138	68.850
7) T	MCPA	6.852	7.343	559.3E6	418.2E6	65.664	69.066
8) T	DICHLORPROP	7.166	7.657	1329.0E6	989.0E6	662.477	681.804
9) T	2,4-D	7.365	7.947	1623.7E6	1188.2E6	665.967	670.297
10) T	Pentachlo...	7.602	8.384	18708.9E6	13808.9E6	681.373	683.613
11) T	2,4,5-TP ...	8.131	8.746	7691.5E6	5859.7E6	682.249	691.592
12) T	2,4,5-T	8.387	9.122	7568.7E6	5595.5E6	680.959	687.699
13) T	2,4-DB	8.898	9.631	1534.4E6	771.4E6	671.787	688.538
14) T	DINOSEB	9.962	9.963	5512.3E6	4127.0E6	673.138	681.037
15) T	Picloram	9.783	10.933	11253.8E6	8090.0E6	675.749	686.297
16) T	DCPA	10.228	10.897	9513.5E6	6990.0E6	683.854	680.374

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS101422\
Data File : PS020798.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Oct 2022 09:33
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 15 00:14:07 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS100622.M
Quant Title : 8080.M
QLast Update : Fri Oct 07 06:09:52 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

