Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS101425\
Data File : PS032019.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Oct 2025 23:57
Operator : AR\AJ
Sample : Q3321-02MS
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
P001-DS-01MS

Manual Integrations APPROVED
Reviewed By :Abdul Mirza 10/22/2025
Supervised By :mohammad ahmed 10/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 15 03:25:11 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS101325.M
Quant Title : 8080.M
QLast Update : Mon Oct 13 12:51:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

		Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
		System Monitoring Compounds						
4)	S	2,4-DCAA	7.278	7.787f	847.6E6	397.5E6	177.810	361.267 #
		Target Compounds						
1)	T	Dalapon	2.651	2.656f	4453.6E6	3389.1E6	576.325	901.770 #
5)	T	DICAMBA	7.463	7.938	3953.6E6	1475.3E6	188.784	170.335m
8)	T	DICHLORPROP	8.172	8.682f	416.5E6	43268264	87.600m	22.095m#
10)	T	Pentachlo...	8.701	9.507	11694.1E6	8056.2E6	174.228	155.453m
11)	T	2,4,5-TP ...	9.280	9.890	7015.3E6	4290.7E6	248.399	213.736m
12)	T	2,4,5-T	9.577	10.323	5524.9E6	3841.7E6	218.858m	224.073m
13)	T	2,4-DB	10.153	10.877	343.1E6	168.3E6	110.679m	130.266m
14)	T	DINOSEB	11.346	11.265	1839.0E6	1039.7E6	82.464	68.389
15)	T	Picloram	11.185	12.391	5265.7E6	8249.7E6	223.331m	247.800
16)	T	DCPA	11.643	12.305	8285.9E6	8398.1E6	225.527	252.865

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

```
Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS101425\
Data File : PS032019.D
Signal(s) : Signal #1: ECD1A.ch   Signal #2: ECD2B.ch
Acq On    : 14 Oct 2025   23:57
Operator  : AR\AJ
Sample    : Q3321-02MS
Misc      :
ALS Vial  : 18   Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 15 03:25:11 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS101325.M
Quant Title  : 8080.M
QLast Update : Mon Oct 13 12:51:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj.      : 1 µl
Signal #1 Phase : ZB-MR1             Signal #2 Phase: ZB-MR2
Signal #1 Info  : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm
```

Instrument :
ECD_S
ClientSampleId :
P001-DS-01MS

Manual Integrations APPROVED

Reviewed By :Abdul Mirza   10/22/2025
Supervised By :mohammad ahmed   10/24/2025