

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS101519\
 Data File : PS007178.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Oct 2019 23:55
 Operator : SM\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 16 01:02:10 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS101419.M
 Quant Title : 8080.M
 QLast Update : Mon Oct 14 13:54:23 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.442	5.175	956.1E6	526.5E6	698.587	748.624
Target Compounds							
1) T	Dalapon	2.020	1.730	1228.0E6	623.7E6	733.894	685.384
2) T	3,5-DICHL...	4.817	4.482	1223.6E6	633.8E6	728.153	736.579
3) T	4-Nitroph...	5.296	4.863	356.2E6	247.2E6	593.121	625.873
5) T	DICAMBA	5.591	5.300	3293.6E6	1859.3E6	670.323	714.549
6) T	MCPD	5.742	5.385	256.0E6	129.5E6	72.621	68.262
7) T	MCPA	5.873	5.560	348.2E6	196.3E6	67.505	67.513
8) T	DICHLORPROP	6.214	5.831	811.5E6	460.7E6	692.305	707.417
9) T	2,4-D	6.429	6.099	873.9E6	566.0E6	709.101	709.407
10) T	Pentachlo...	6.685	6.462	13077.3E6	7237.3E6	732.534	752.597
11) T	2,4,5-TP ...	7.218	6.824	4959.8E6	2794.1E6	711.621	734.590
12) T	2,4,5-T	7.490	7.182	5309.0E6	2626.2E6	694.465	749.577
13) T	2,4-DB	8.018	7.667	631.6E6	350.7E6	662.461	748.340
14) T	DINOSEB	9.099	7.975	4017.8E6	2127.7E6	690.443	757.908
15) T	Picloram	8.934	8.910	6931.5E6	3916.3E6	628.009	637.810
16) T	DCPA	9.372	8.867	6519.4E6	3792.1E6	728.076	756.209

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS101519\
Data File : PS007178.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Oct 2019 23:55
Operator : SM\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 16 01:02:10 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS101419.M
Quant Title : 8080.M
QLast Update : Mon Oct 14 13:54:23 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

