

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS101520\
 Data File : PS012863.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Oct 2020 20:41
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 16 02:05:23 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS092220.M
 Quant Title : 8080.M
 QLast Update : Wed Sep 23 04:57:25 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.213	6.910	732.9E6	331.9E6	748.272	721.591
Target Compounds							
1) T	Dalapon	3.033	2.625	726.4E6	369.4E6	639.203	617.540
2) T	3,5-DICHL...	6.429	6.064	936.1E6	482.7E6	694.217	677.004
3) T	4-Nitroph...	7.047	6.522	323.9E6	180.1E6	666.078	644.748
5) T	DICAMBA	7.403	7.077	2747.1E6	1324.2E6	712.121	694.432
6) T	MCPD	7.571	7.172	210.5E6	104.7E6	72.156	69.205
7) T	MCPA	7.720	7.388	294.3E6	163.1E6	70.560	68.031
8) T	DICHLORPROP	8.095	7.703	681.4E6	353.5E6	704.141	684.553
9) T	2,4-D	8.326	8.000	792.3E6	427.1E6	697.775	681.938
10) T	Pentachlo...	8.634	8.425	10029.4E6	4386.1E6	725.724	719.112
11) T	2,4,5-TP ...	9.182	8.787	3971.6E6	1873.6E6	716.267	710.120
12) T	2,4,5-T	9.470	9.167	3995.3E6	1798.7E6	710.760	699.552
13) T	2,4-DB	10.024	9.674	644.2E6	244.6E6	680.706	653.248
14) T	DINOSEB	11.178	10.003	3267.2E6	1356.0E6	719.607	707.932
15) T	Picloram	11.011	10.997	4601.9E6	2531.8E6	639.141	642.359
16) T	DCPA	11.444	10.931	5109.5E6	2408.9E6	723.392	724.403

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS101520\
Data File : PS012863.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Oct 2020 20:41
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 16 02:05:23 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS092220.M
Quant Title : 8080.M
QLast Update : Wed Sep 23 04:57:25 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

