

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS101620\
 Data File : PS012884.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Oct 2020 21:12
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 17 01:55:39 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS092220.M
 Quant Title : 8080.M
 QLast Update : Wed Sep 23 04:57:25 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.213	6.910	752.1E6	339.0E6	767.868	736.878
Target Compounds							
1) T	Dalapon	3.033	2.624	683.8E6	350.8E6	601.757	586.439
2) T	3,5-DICHL...	6.429	6.064	963.7E6	493.3E6	714.652	691.913
3) T	4-Nitroph...	7.048	6.522	327.9E6	181.8E6	674.262	650.484
5) T	DICAMBA	7.403	7.077	2835.7E6	1357.5E6	735.074	711.903
6) T	MCPD	7.571	7.172	219.8E6	107.6E6	75.327	71.135
7) T	MCPA	7.720	7.388	306.2E6	164.9E6	73.427	68.779
8) T	DICHLORPROP	8.095	7.703	701.6E6	360.4E6	725.068	697.948
9) T	2,4-D	8.327	8.000	803.1E6	433.7E6	707.305	692.550
10) T	Pentachlo...	8.634	8.426	10126.6E6	4438.0E6	732.761	727.616
11) T	2,4,5-TP ...	9.183	8.787	4073.4E6	1905.4E6	734.634	722.181
12) T	2,4,5-T	9.470	9.167	4048.9E6	1822.5E6	720.280	708.797
13) T	2,4-DB	10.024	9.674	641.9E6	273.5E6	678.350	730.672
14) T	DINOSEB	11.177	10.002	3352.3E6	1383.1E6	738.358	722.103
15) T	Picloram	11.012	10.998	4439.3E6	2466.8E6	616.559	625.862
16) T	DCPA	11.443	10.930	5241.3E6	2463.0E6	742.057	740.664

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS101620\
Data File : PS012884.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Oct 2020 21:12
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 17 01:55:39 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS092220.M
Quant Title : 8080.M
QLast Update : Wed Sep 23 04:57:25 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

