

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS101821\
 Data File : PS017105.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Oct 2021 19:07
 Operator : DD\AJ
 Sample : HSTDICV750
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 19 01:23:08 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS101821.M
 Quant Title : 8080.M
 QLast Update : Tue Oct 19 01:21:46 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 T	Dalapon	2.171	2.369 E6	-9.1	105	0.00
2 T	3,5-DICHLORO BENZOIC ACID	2.159	2.274 E6	-5.3	105	0.00
3 T	4-Nitrophenol	872.156	931.084 E3	-6.8	105	0.00
4 S	2,4-DCAA	1.528	1.609 E6	-5.3	105	0.00
5 T	DICAMBA	5.990	6.372 E6	-6.4	105	0.00
6 T	MCP	4.470	4.914 E6	-9.9	106	0.00
7 T	MCPA	6.169	6.718 E6	-8.9	106	0.00
8 T	DICHLORPROP	1.499	1.582 E6	-5.5	105	0.00
9 T	2,4-D	1.771	1.912 E6	-8.0	107	0.00
10 T	Pentachlorophenol	20.748	21.949 E6	-5.8	104	0.00
11 T	2,4,5-TP (SILVEX)	8.133	8.624 E6	-6.0	105	0.00
12 T	2,4,5-T	8.032	8.607 E6	-7.2	105	0.00
13 T	2,4-DB	1.504	1.587 E6	-5.5	104	0.00
14 T	DINOSEB	5.665	6.154 E6	-8.6	106	0.00
15 T	Picloram	10.681	11.722 E6	-9.7	106	0.00
16 T	DCPA	10.184	10.839 E6	-6.4	105	0.00

Signal #2

1 T	Dalapon	647.044	694.495 E3	-7.3	105	0.00
2 T	3,5-DICHLORO BENZOIC ACID	868.532	931.508 E3	-7.3	105	0.00
3 T	4-Nitrophenol	322.345	337.631 E3	-4.7	105	0.00
4 S	2,4-DCAA	584.442	621.571 E3	-6.4	105	0.00
5 T	DICAMBA	2.546	2.724 E6	-7.0	105	0.00
6 T	MCP	2.014	2.187 E6	-8.6	106	0.00
7 T	MCPA	2.960	3.147 E6	-6.3	106	0.00
8 T	DICHLORPROP	643.094	686.546 E3	-6.8	106	0.00
9 T	2,4-D	752.581	811.541 E3	-7.8	106	0.00
10 T	Pentachlorophenol	8.165	8.650 E6	-5.9	105	0.00
11 T	2,4,5-TP (SILVEX)	3.447	3.699 E6	-7.3	105	0.00
12 T	2,4,5-T	3.169	3.426 E6	-8.1	105	0.00
13 T	2,4-DB	414.776	443.331 E3	-6.9	105	0.00
14 T	DINOSEB	2.207	2.445 E6	-10.8	107	0.00
15 T	Picloram	3.830	4.239 E6	-10.7	106	0.00
16 T	DCPA	4.267	4.544 E6	-6.5	105	0.00

Evaluate Continuing Calibration Report - Not Found

Signal #2

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS101821\
Data File : PS017105.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Oct 2021 19:07
Operator : DD\AJ
Sample : HSTDICV750
Misc :
ALS Vial : 8 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 19 01:23:08 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS101821.M
Quant Title : 8080.M
QLast Update : Tue Oct 19 01:21:46 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0