

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS101921\
 Data File : PS017127.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Oct 2021 18:34
 Operator : DD\AJ
 Sample : M4194-03MS
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 VNJ-209MS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 20 01:34:00 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS101821.M
 Quant Title : 8080.M
 QLast Update : Tue Oct 19 01:21:46 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.153	6.738	938.8E6	395.6E6	614.242	676.849m
Target Compounds							
1) T	Dalapon	2.987	2.495	801.3E6	259.7E6	369.029	401.308m
2) T	3,5-DICHL...	6.377	5.936	1126.0E6	462.4E6	521.595	532.370
3) T	4-Nitroph...	6.982	6.363	401.9E6	154.6E6	460.823	479.654
5) T	DICAMBA	7.343	6.900	2920.5E6	1271.7E6	487.585	499.460
6) T	MCPD	7.511	6.994	205.1E6	91402000	45.880	45.386
7) T	MCPA	7.658	7.204	282.2E6	145.3E6	45.740	49.087
8) T	DICHLORPROP	8.034	7.517	655.5E6	281.1E6	437.193	437.102
9) T	2,4-D	8.262	7.807	879.9E6	405.4E6	496.787	538.683
10) T	Pentachlo...	8.568	8.231	7998.2E6	3217.8E6	385.491	394.109
11) T	2,4,5-TP ...	9.117	8.592	3699.6E6	1591.2E6	454.908	461.592m
12) T	2,4,5-T	9.401	8.970	3502.1E6	1479.6E6	436.008	466.832
13) T	2,4-DB	9.952	9.474	485.8E6	179.6E6	322.932	433.017 #
14) T	DINOSEB	11.103	9.800	480.5E6	180.5E6	84.813	81.783
15) T	Picloram	10.925	10.778	5992.4E6	2193.1E6	561.014	572.658
16) T	DCPA	11.372	10.729	4218.9E6	2228.0E6	414.256	522.106 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS101921\
Data File : PS017127.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Oct 2021 18:34
Operator : DD\AJ
Sample : M4194-03MS
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
VNJ-209MS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 20 01:34:00 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS101821.M
Quant Title : 8080.M
QLast Update : Tue Oct 19 01:21:46 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

