

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS101921\
 Data File : PS017131.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Oct 2021 20:11
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 20 01:35:16 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS101821.M
 Quant Title : 8080.M
 QLast Update : Tue Oct 19 01:21:46 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.151	6.737	1174.1E6	459.7E6	768.186	786.481
Target Compounds							
1) T	Dalapon	2.990	2.493	1579.6E6	483.6E6	727.475	747.347
2) T	3,5-DICHL...	6.375	5.935	1546.6E6	641.9E6	716.435	739.035
3) T	4-Nitroph...	6.981	6.362	614.1E6	233.2E6	704.138	723.345
5) T	DICAMBA	7.340	6.898	4383.1E6	1898.8E6	731.769	745.797
6) T	MCPD	7.511	6.995	334.6E6	152.5E6	74.849	75.746
7) T	MCPA	7.658	7.206	463.5E6	214.8E6	75.126	72.563
8) T	DICHLORPROP	8.031	7.515	1089.1E6	479.9E6	726.360	746.188
9) T	2,4-D	8.259	7.805	1321.7E6	565.8E6	746.205	751.780
10) T	Pentachlo...	8.564	8.230	15226.3E6	6112.1E6	733.868	748.602
11) T	2,4,5-TP ...	9.114	8.590	6011.4E6	2609.8E6	739.164	757.083
12) T	2,4,5-T	9.398	8.967	6069.4E6	2454.8E6	755.635	774.502
13) T	2,4-DB	9.948	9.472	1124.6E6	312.4E6	747.540	753.169
14) T	DINOSEB	11.100	9.797	4367.8E6	1760.5E6	771.044	797.853
15) T	Picloram	10.922	10.775	8687.8E6	3189.3E6	813.358	832.766
16) T	DCPA	11.368	10.727	7663.5E6	3245.1E6	752.486	760.454

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS101921\
Data File : PS017131.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Oct 2021 20:11
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 20 01:35:16 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS101821.M
Quant Title : 8080.M
QLast Update : Tue Oct 19 01:21:46 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

