

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\
 Method File : PS102020.M
 Title : 8080.M
 Last Update : Wed Oct 21 02:27:47 2020
 Response Via : Initial Calibration

Calibration Files

750 =PS012908.D 1000 =PS012909.D 1500 =PS012910.D
 500 =PS012907.D 200 =PS012906.D

	Compound	750	1000	1500	500	200	Avg		%RSD
1) T	Dalapon	1.000	1.031	1.084	1.047	0.921	1.017	E6	6.01
2) T	3,5-DICHLOROBENZOI...	1.210	1.228	1.261	1.289	1.203	1.238	E6	2.93
3) T	4-Nitrophenol	4.162	4.309	4.630	4.289	3.521	4.182	E5	9.75
4) S	2,4-DCAA	8.774	8.957	9.275	9.264	8.484	8.951	E5	3.76
5) T	DICAMBA	3.516	3.560	3.615	3.741	3.372	3.561	E6	3.80
6) T	MCPP	2.655	2.771	2.997	2.705	2.188	2.663	E6	11.12
7) T	MCPA	3.769	3.920	4.167	3.923	3.366	3.829	E6	7.72
8) T	DICHLORPROP	8.708	8.868	9.192	9.261	8.565	8.919	E5	3.38
9) T	2,4-D	1.014	1.035	1.080	1.073	0.958	1.032	E6	4.81
10) T	Pentachlorophenol	1.265	1.242	1.190	1.393	1.329	1.284	E7	6.12
11) T	2,4,5-TP (SILVEX)	5.046	5.056	5.075	5.440	4.992	5.122	E6	3.52
12) T	2,4,5-T	5.172	5.158	5.200	5.555	5.097	5.237	E6	3.47
13) T	2,4-DB	8.531	8.920	9.684	8.647	6.425	8.441	E5	14.37
14) T	DINOSEB	4.202	4.166	4.183	4.501	4.384	4.287	E6	3.46
15) T	Picloram	6.465	6.700	7.200	6.674	4.610	6.330	E6	15.77
16) T	DCPA	6.502	6.446	6.391	7.090	6.703	6.626	E6	4.30

Signal #2 Calibration Files

750 =PS012908.D 1000 =PS012909.D 1500 =PS012910.D
 500 =PS012907.D 200 =PS012906.D

	Compound	750	1000	1500	500	200	Avg		%RSD
1) T	Dalapon	5.175	5.248	5.401	5.497	4.972	5.259	E5	3.88
2) T	3,5-DICHLOROBENZOI...	6.129	6.293	6.301	6.556	6.226	6.301	E5	2.51
3) T	4-Nitrophenol	2.258	2.353	2.473	2.321	2.227	2.326	E5	4.11
4) S	2,4-DCAA	3.899	3.991	4.119	4.133	3.920	4.013	E5	2.73
5) T	DICAMBA	1.678	1.684	1.671	1.803	1.673	1.702	E6	3.35
6) T	MCPP	1.317	1.356	1.434	1.372	1.176	1.331	E6	7.25
7) T	MCPA	2.095	2.125	2.183	2.260	2.291	2.191	E6	3.84
8) T	DICHLORPROP	4.426	4.513	4.575	4.783	4.672	4.594	E5	3.02
9) T	2,4-D	5.403	5.497	5.631	5.758	5.199	5.498	E5	3.89
10) T	Pentachlorophenol	5.542	5.349	4.940	6.211	6.199	5.648	E6	9.79
11) T	2,4,5-TP (SILVEX)	2.328	2.315	2.264	2.543	2.403	2.371	E6	4.57
12) T	2,4,5-T	2.246	2.257	2.249	2.413	2.136	2.260	E6	4.37
13) T	2,4-DB	3.307	3.340	3.444	3.342	3.180	3.322	E5	2.86
14) T	DINOSEB	1.669	1.661	1.674	1.755	1.649	1.682	E6	2.50
15) T	Picloram	3.522	3.611	3.563	3.942	3.361	3.600	E6	5.92
16) T	DCPA	2.973	2.929	2.817	3.300	3.218	3.047	E6	6.67

(#) = Out of Range