

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS102020\
 Data File : PS012911.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Oct 2020 19:47
 Operator : DD\AJ
 Sample : HSTDICV750
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 ICVPS102020

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 02:30:41 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS102020.M
 Quant Title : 8080.M
 QLast Update : Wed Oct 21 02:27:47 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 T	Dalapon	1.017	1.038 E6	-2.1	104	0.00
2 T	3,5-DICHLOROBENZOIC ACID	1.238	1.258 E6	-1.6	104	0.00
3 T	4-Nitrophenol	418.215	431.926 E3	-3.3	104	0.00
4 S	2,4-DCAA	895.069	909.828 E3	-1.6	104	0.00
5 T	DICAMBA	3.561	3.653 E6	-2.6	104	0.00
6 T	MCPD	2.663	2.767 E6	-3.9	104	0.00
7 T	MCPA	3.829	3.929 E6	-2.6	104	0.00
8 T	DICHLOROPROP	891.870	905.560 E3	-1.5	104	0.00
9 T	2,4-D	1.032	1.051 E6	-1.8	104	0.00
10 T	Pentachlorophenol	12.838	13.097 E6	-2.0	104	0.00
11 T	2,4,5-TP (SILVEX)	5.122	5.235 E6	-2.2	104	0.00
12 T	2,4,5-T	5.237	5.339 E6	-1.9	103	0.00
13 T	2,4-DB	844.142	891.650 E3	-5.6	105	0.00
14 T	DINOSEB	4.287	4.294 E6	-0.2	102	0.00
15 T	Picloram	6.330	6.794 E6	-7.3	105	0.00
16 T	DCPA	6.626	6.716 E6	-1.4	103	0.00

Signal #2

1 T	Dalapon	525.876	533.299 E3	-1.4	103	0.00
2 T	3,5-DICHLOROBENZOIC ACID	630.100	639.911 E3	-1.6	104	0.00
3 T	4-Nitrophenol	232.636	234.897 E3	-1.0	104	0.00
4 S	2,4-DCAA	401.252	407.368 E3	-1.5	104	0.00
5 T	DICAMBA	1.702	1.749 E6	-2.8	104	0.00
6 T	MCPD	1.331	1.379 E6	-3.6	105	0.00
7 T	MCPA	2.191	2.181 E6	0.5	104	0.00
8 T	DICHLOROPROP	459.370	463.326 E3	-0.9	105	0.00
9 T	2,4-D	549.784	562.019 E3	-2.2	104	0.00
10 T	Pentachlorophenol	5.648	5.748 E6	-1.8	104	0.00
11 T	2,4,5-TP (SILVEX)	2.371	2.419 E6	-2.0	104	0.00
12 T	2,4,5-T	2.260	2.343 E6	-3.7	104	0.00
13 T	2,4-DB	332.248	338.672 E3	-1.9	102	0.00
14 T	DINOSEB	1.682	1.721 E6	-2.3	103	0.00
15 T	Picloram	3.600	3.716 E6	-3.2	105	0.00
16 T	DCPA	3.047	3.091 E6	-1.4	104	0.00

Evaluate Continuing Calibration Report - Not Found

Signal #2

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS102020\
Data File : PS012911.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Oct 2020 19:47
Operator : DD\AJ
Sample : HSTDICV750
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
ICVPS102020

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 02:30:41 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS102020.M
Quant Title : 8080.M
QLast Update : Wed Oct 21 02:27:47 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
----------	-------	------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0