

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS102020\
 Data File : PS012922.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2020 00:01
 Operator : DD\AJ
 Sample : L4214-08
 Misc : LOQ-WATER-
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 LOQ-WATER-02-QT4-2020

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 02:49:56 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS102020.M
 Quant Title : 8080.M
 QLast Update : Wed Oct 21 02:27:47 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.212	6.914	466.6E6	246.6E6	521.257	614.459
Target Compounds							
1) T	Dalapon	3.027	2.617	217.6E6	221.4E6	214.046	421.049 #
2) T	3,5-DICHL...	6.429	6.070	230.9E6	124.8E6	186.435	197.989
3) T	4-Nitroph...	7.052	6.534	65866071	28431433	157.493	122.214
5) T	DICAMBA	7.402	7.080	623.5E6	318.0E6	175.091	186.884
6) T	MCPD	7.565	7.169	54484808	23350743	20.459	17.543
7) T	MCPA	7.713	7.381	70003660	65083399	18.282	29.708 #
8) T	DICHLORPROP	8.093	7.707	167.2E6	93244860	187.496	202.984
9) T	2,4-D	8.329	8.008	203.6E6	106.8E6	197.289	194.248
10) T	Pentachlo...	8.631	8.428	2640.6E6	1234.7E6	205.685	218.590
11) T	2,4,5-TP ...	9.181	8.789	1004.2E6	511.7E6	196.066	215.837
12) T	2,4,5-T	9.470	9.174	975.9E6	427.8E6	186.374	189.276
13) T	2,4-DB	10.025	9.681	106.5E6	60261956	126.132	181.376 #
14) T	DINOSEB	11.174	10.005	733.8E6	299.9E6	171.155	178.360
15) T	Picloram	11.024	11.013	840.7E6	553.5E6	132.812	153.754
16) T	DCPA	11.441	10.931	1140.4E6	617.5E6	172.099	202.630

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS102020\
Data File : PS012922.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2020 00:01
Operator : DD\AJ
Sample : L4214-08
Misc : LOQ-WATER-
ALS Vial : 29 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
LOQ-WATER-02-QT4-2020

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 02:49:56 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS102020.M
Quant Title : 8080.M
QLast Update : Wed Oct 21 02:27:47 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

