

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS102119\
 Data File : PS007244.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2019 21:19
 Operator : SM\AJ
 Sample : PB124159BL
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 22 05:15:10 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS101419.M
 Quant Title : 8080.M
 QLast Update : Mon Oct 14 13:54:23 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.475	5.184	851.0E6	474.4E6	621.803	674.590
Target Compounds							
1) T	Dalapon	2.032	1.721	97603642	221.1E6	58.331	242.950 #
2) T	3,5-DICHL...	0.000	4.518	0	31690070	N.D.	36.830
5) T	DICAMBA	5.569	0.000	52118494	0	10.607	N.D. #
7) T	MCPA	5.881	5.556	75814832	21559818	14.699	7.416 #
8) T	DICHLORPROP	0.000	5.775	0	50890479	N.D.	78.147
9) T	2,4-D	0.000	6.092	0	9417155	N.D.	11.804
10) T	Pentachlo...	0.000	6.477	0	14859555	N.D.	1.545
11) T	2,4,5-TP ...	0.000	6.786	0	23837236	N.D.	6.267
12) T	2,4,5-T	7.437	0.000	93779158	0	12.267	N.D. #
14) T	DINOSEB	0.000	7.978	0	9322020	N.D.	3.321
15) T	Picloram	0.000	8.942	0	13848655	N.D.	2.255



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS102119\
Data File : PS007244.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2019 21:19
Operator : SM\AJ
Sample : PB124159BL
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 22 05:15:10 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS101419.M
Quant Title : 8080.M
QLast Update : Mon Oct 14 13:54:23 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 2.032 min
Delta R.T.: 0.021 min
Response: 97603642
Conc: 58.33 ng/ml

Instrument :
ECD_S
ClientSampleId :




#1 Dalapon

R.T.: 1.721 min
Delta R.T.: 0.000 min
Response: 221085086
Conc: 242.95 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 0.000 min
Exp R.T. : 4.810 min
Response: 0
Conc: N.D.



#2 3,5-DICHLOROBENZOIC ACID


R.T.: 4.518 min
Delta R.T.: 0.043 min
Response: 31690070
Conc: 36.83 ng/ml



#4 2,4-DCAA


R.T.: 5.475 min
Delta R.T.: 0.039 min
Response: 851042900
Conc: 621.80 ng/ml

Instrument :
ECD_S
ClientSampleId :




#4 2,4-DCAA

R.T.: 5.184 min
Delta R.T.: 0.015 min
Response: 474443745
Conc: 674.59 ng/ml




#5 DICAMBA

R.T.: 5.569 min
Delta R.T.: -0.016 min
Response: 52118494
Conc: 10.61 ng/ml



#5 DICAMBA


R.T.: 0.000 min
Exp R.T. : 5.294 min
Response: 0
Conc: N.D.



#7 MCPA


R.T.: 5.881 min
Delta R.T.: 0.014 min
Response: 75814832
Conc: 14.70 ug/ml

Instrument :
ECD_S
ClientSampleId :




#7 MCPA

R.T.: 5.556 min
Delta R.T.: 0.000 min
Response: 21559818
Conc: 7.42 ug/ml




#8 DICHLORPROP

R.T.: 0.000 min
Exp R.T. : 6.209 min
Response: 0
Conc: N.D.



#8 DICHLORPROP


R.T.: 5.775 min
Delta R.T.: -0.052 min
Response: 50890479
Conc: 78.15 ng/ml



#9 2,4-D


R.T.: 0.000 min
Exp R.T.: 6.424 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :




#9 2,4-D

R.T.: 6.092 min
Delta R.T.: -0.004 min
Response: 9417155
Conc: 11.80 ng/ml




#10 Pentachlorophenol

R.T.: 0.000 min
Exp R.T.: 6.680 min
Response: 0
Conc: N.D.



#10 Pentachlorophenol


R.T.: 6.477 min
Delta R.T.: 0.019 min
Response: 14859555
Conc: 1.55 ng/ml



#11 2,4,5-TP (SILVEX)


R.T.: 0.000 min
Exp R.T.: 7.214 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :




#11 2,4,5-TP (SILVEX)

R.T.: 6.786 min
Delta R.T.: -0.034 min
Response: 23837236
Conc: 6.27 ng/ml




#12 2,4,5-T

R.T.: 7.437 min
Delta R.T.: -0.049 min
Response: 93779158
Conc: 12.27 ng/ml



#12 2,4,5-T


R.T.: 0.000 min
Exp R.T.: 7.178 min
Response: 0
Conc: N.D.



#14 DINOSEB


R.T.: 0.000 min
Exp R.T. : 9.095 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :




#14 DINOSEB

R.T.: 7.978 min
Delta R.T.: 0.007 min
Response: 9322020
Conc: 3.32 ng/ml



#15 Picloram

R.T.: 0.000 min
Exp R.T. : 8.929 min
Response: 0
Conc: N.D.



#15 Picloram

R.T.: 8.942 min
Delta R.T.: 0.038 min
Response: 13848655
Conc: 2.26 ng/ml