

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS102119\
 Data File : PS007245.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2019 21:41
 Operator : SM\AJ
 Sample : PB124159BS
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleID :
 PB124159BS

Manual Integrations
 APPROVED

Ankita
 10/22/2019 12:27:08 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 22 05:15:23 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS101419.M
 Quant Title : 8080.M
 QLast Update : Mon Oct 14 13:54:23 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.434	5.165	931.7E6	516.7E6	680.707	734.716
Target Compounds							
1) T	Dalapon	2.011	1.724	1069.7E6	654.2E6	639.304	718.879
2) T	3,5-DICHL...	4.808	4.472	1065.3E6	535.3E6	633.961	622.163m
3) T	4-Nitroph...	5.287	4.854	353.2E6	194.2E6	588.183	491.758
5) T	DICAMBA	5.584	5.290	2880.5E6	1535.9E6	586.246	590.258
6) T	MCPD	5.735	5.375	234.1E6	118.4E6	66.413m	62.431m
7) T	MCPA	5.866	5.549	338.6E6	182.6E6	65.656	62.795
8) T	DICHLORPROP	6.207	5.820	766.2E6	376.0E6	653.592	577.387
9) T	2,4-D	6.419	6.087	945.2E6	574.5E6	766.956	720.075
10) T	Pentachlo...	6.678	6.450	11873.0E6	6301.8E6	665.076	655.314
11) T	2,4,5-TP ...	7.212	6.811	4720.3E6	2550.0E6	677.265	670.407
12) T	2,4,5-T	7.481	7.170	4699.7E6	2338.2E6	614.765	667.355
13) T	2,4-DB	8.009	7.654	562.0E6	305.8E6	589.408	652.397
14) T	DINOSEB	9.093	7.962	3467.1E6	1784.5E6	595.806	635.658
15) T	Picloram	8.925	8.898	6021.7E6	3194.4E6	545.576	520.249m
16) T	DCPA	9.365	8.857	4079.0E6	1928.0E6	455.537	384.468m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS102119\
Data File : PS007245.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2019 21:41
Operator : SM\AJ
Sample : PB124159BS
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
PB124159BS

Manual Integrations
APPROVED

Ankita
10/22/2019 12:27:08 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 22 05:15:23 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS101419.M
Quant Title : 8080.M
QLast Update : Mon Oct 14 13:54:23 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

