

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS102218\
 Data File : PS002071.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Oct 2018 12:14
 Operator : EI\SJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Sohil
 10/23/2018 12:33:30 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 23 05:10:50 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS101518.M
 Quant Title : 8080.M
 QLast Update : Wed Oct 17 08:59:14 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.446	10.940	146.4E6	176.2E6	692.551	738.555
Target Compounds							
1) T	Dalapon	2.927	2.390	192.3E6	191.2E6	630.166	654.801
2) T	3,5-DICHL...	9.869	9.091	199.6E6	201.2E6	656.975	688.016
3) T	4-Nitroph...	11.036	10.085	89699165	85131652	664.758	692.370
5) T	DICAMBA	11.758	11.224	525.3E6	570.1E6	667.338	693.832
6) T	MCPD	12.096	11.451	38857021	39622442	69.394m	72.598
7) T	MCPA	12.347	11.826	62662346	59279581	67.793m	72.005
8) T	DICHLORPROP	12.948	12.349	143.1E6	151.4E6	660.492	691.015
9) T	2,4-D	13.307	12.821	179.3E6	204.0E6	671.290	682.823
10) T	Pentachlo...	13.687	13.369	1945.1E6	2248.1E6	687.328	710.004
11) T	2,4,5-TP ...	14.559	13.987	743.9E6	893.8E6	680.080	703.566
12) T	2,4,5-T	14.957	14.554	783.3E6	748.9E6	667.889	683.714
13) T	2,4-DB	15.706	15.203	106.3E6	113.8E6	676.673	722.968
14) T	DINOSB	17.191	15.632	544.5E6	608.7E6	650.337	688.281
15) T	Picloram	16.956	16.812	1154.8E6	1121.2E6	675.459	706.802
16) T	DCPA	17.534	16.866	956.6E6	1285.7E6	690.116	710.634

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS102218\
Data File : PS002071.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Oct 2018 12:14
Operator : EI\SJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
HSTDCCC750

Manual Integrations
APPROVED

Sohil
10/23/2018 12:33:30 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 23 05:10:50 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS101518.M
Quant Title : 8080.M
QLast Update : Wed Oct 17 08:59:14 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

