

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS102220\
 Data File : PS012961.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Oct 2020 20:14
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 23 02:29:31 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS102020.M
 Quant Title : 8080.M
 QLast Update : Wed Oct 21 02:27:47 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.211	6.914	680.1E6	308.9E6	759.779	769.925
Target Compounds							
1) T	Dalapon	3.032	2.623	679.8E6	356.4E6	668.755	677.666
2) T	3,5-DICHL...	6.427	6.067	869.3E6	453.8E6	701.965	720.126
3) T	4-Nitroph...	7.045	6.526	301.3E6	177.0E6	720.399	760.962
5) T	DICAMBA	7.401	7.080	2566.2E6	1241.5E6	720.699	729.480
6) T	MCPD	7.567	7.174	193.5E6	100.0E6	72.669	75.140
7) T	MCPA	7.716	7.389	274.6E6	154.9E6	71.706	70.716
8) T	DICHLORPROP	8.091	7.705	635.9E6	330.1E6	712.940	718.621
9) T	2,4-D	8.323	8.003	736.7E6	400.3E6	713.980	728.027
10) T	Pentachlo...	8.630	8.427	9290.9E6	4150.8E6	723.699	734.884
11) T	2,4,5-TP ...	9.178	8.789	3701.8E6	1748.1E6	722.744	737.422
12) T	2,4,5-T	9.465	9.169	3758.2E6	1675.0E6	717.701	741.060
13) T	2,4-DB	10.020	9.676	633.5E6	229.7E6	750.432	691.346
14) T	DINOSEB	11.171	10.003	3043.4E6	1245.6E6	709.882	740.741
15) T	Picloram	11.007	10.999	4487.8E6	2574.5E6	708.964	715.193
16) T	DCPA	11.437	10.931	4763.4E6	2314.5E6	718.860	759.502

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS102220\
Data File : PS012961.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Oct 2020 20:14
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 23 02:29:31 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS102020.M
Quant Title : 8080.M
QLast Update : Wed Oct 21 02:27:47 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

