

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS102318\
 Data File : PS002080.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Oct 2018 12:01
 Operator : EI\SJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Sohil
 10/24/2018 3:45:09 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 24 00:59:19 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS101518.M
 Quant Title : 8080.M
 QLast Update : Wed Oct 17 08:59:14 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.446	10.938	151.9E6	179.0E6	718.637	750.542
Target Compounds							
1) T	Dalapon	2.926	2.390	194.8E6	192.2E6	638.273	658.266
2) T	3,5-DICHL...	9.869	9.090	204.1E6	204.2E6	671.724	698.204
3) T	4-Nitroph...	11.035	10.084	92023679	85604961	681.984	696.220
5) T	DICAMBA	11.757	11.222	537.1E6	581.1E6	682.372	707.248
6) T	MCPD	12.096	11.449	39535291	40116559	70.605m	73.503
7) T	MCPA	12.347	11.825	63217689	59295193	68.393m	72.023
8) T	DICHLORPROP	12.948	12.347	147.2E6	154.5E6	679.448	705.402
9) T	2,4-D	13.306	12.819	184.7E6	209.1E6	691.650	699.763
10) T	Pentachlo...	13.686	13.367	1981.1E6	2277.5E6	700.051	719.298
11) T	2,4,5-TP ...	14.558	13.985	758.1E6	907.5E6	693.047	714.365
12) T	2,4,5-T	14.956	14.552	797.4E6	758.3E6	679.931	692.305
13) T	2,4-DB	15.704	15.201	109.2E6	116.7E6	694.955	741.480
14) T	DINOSEB	17.191	15.631	559.4E6	623.3E6	668.141	704.763
15) T	Picloram	16.955	16.810	1196.8E6	1142.4E6	700.000	720.221
16) T	DCPA	17.533	16.864	973.7E6	1325.3E6	702.450	732.527

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS102318\
Data File : PS002080.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Oct 2018 12:01
Operator : EI\SJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Sohil
10/24/2018 3:45:09 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 24 00:59:19 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS101518.M
Quant Title : 8080.M
QLast Update : Wed Oct 17 08:59:14 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

