

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS102324\
 Data File : PS028007.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Oct 2024 11:04
 Operator : AR\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 23 13:27:17 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS102324.M
 Quant Title : 8080.M
 QLast Update : Wed Oct 23 13:25:49 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.091	7.616	1317.6E6	437.0E6	488.565	460.591
Target Compounds							
1) T	Dalapon	2.563	0.000	13020003	0	2.858	N.D. #
3) T	4-Nitroph...	6.890	0.000	7620379	0	4.118	N.D. #
10) T	Pentachlo...	8.434f	0.000	1781478	0	<MDL	N.D. #
12) T	2,4,5-T	9.354f	0.000	1418606	0	<MDL	N.D. #
13) T	2,4-DB	9.828f	0.000	4677593	0	1.777	N.D. #
14) T	DINOSEB	11.024f	0.000	3716340	0	<MDL	N.D. #
15) T	Picloram	10.877	12.161	37306987	3540211	1.676	<MDL #
16) T	DCPA	11.339	0.000	1665786	0	<MDL	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS102324\
Data File : PS028007.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Oct 2024 11:04
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 23 13:27:17 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS102324.M
Quant Title : 8080.M
QLast Update : Wed Oct 23 13:25:49 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 2.563 min
Delta R.T.: 0.025 min
Response: 13020003
Conc: 2.86 ng/ml

Instrument :
ECD_S
ClientSampleId :
I.BLK




#1 Dalapon

R.T.: 0.000 min
Exp R.T. : 2.613 min
Response: 0
Conc: N.D.




#3 4-Nitrophenol

R.T.: 6.890 min
Delta R.T.: 0.003 min
Response: 7620379
Conc: 4.12 ng/ml



#3 4-Nitrophenol


R.T.: 0.000 min
Exp R.T. : 7.149 min
Response: 0
Conc: N.D.



#4 2,4-DCAA


R.T.: 7.091 min
Delta R.T.: 0.000 min
Response: 1317642979
Conc: 488.57 ng/ml

Instrument :
ECD_S
ClientSampleId :
I.BLK




#4 2,4-DCAA

R.T.: 7.616 min
Delta R.T.: 0.001 min
Response: 436952671
Conc: 460.59 ng/ml




#13 2,4-DB

R.T.: 9.828 min
Delta R.T.: -0.049 min
Response: 4677593
Conc: 1.78 ng/ml



#13 2,4-DB


R.T.: 0.000 min
Exp R.T. : 10.702 min
Response: 0
Conc: N.D.



#15 Picloram

R.T.: 10.877 min
Delta R.T.: 0.010 min
Response: 37306987
Conc: 1.68 ng/ml

Instrument :
ECD_S
ClientSampleId :
I.BLK



#15 Picloram

R.T.: 12.161 min
Delta R.T.: 0.008 min
Response: 3540211
Conc: N.D.