

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS102419\
 Data File : PS007308.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Oct 2019 19:59
 Operator : SM\AJ
 Sample : K5497-11
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 25 01:52:48 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS101419.M
 Quant Title : 8080.M
 QLast Update : Mon Oct 14 13:54:23 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.453	5.182	308.3E6	159.1E6	225.237	226.280
Target Compounds							
1) T	Dalapon	2.011	1.720	94129676	371.9E6	56.255	408.647 #
2) T	3,5-DICHL...	0.000	4.515	0	45787051	N.D.	53.214
5) T	DICAMBA	5.547	0.000	42616082	0	8.673	N.D. #
7) T	MCPA	0.000	5.552	0	23791627	N.D.	8.184
9) T	2,4-D	0.000	6.086	0	11847480	N.D.	14.850
10) T	Pentachlo...	0.000	6.475	0	9699009	N.D.	1.009
11) T	2,4,5-TP ...	0.000	6.780	0	58890156	N.D.	15.483
12) T	2,4,5-T	7.422	0.000	150.0E6	0	19.626	N.D. #
13) T	2,4-DB	0.000	7.680	0	18768368	N.D.	40.044
14) T	DINOSEB	0.000	8.009	0	14563877	N.D.	5.188



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS102419\
Data File : PS007308.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Oct 2019 19:59
Operator : SM\AJ
Sample : K5497-11
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 25 01:52:48 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS101419.M
Quant Title : 8080.M
QLast Update : Mon Oct 14 13:54:23 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 2.011 min
Delta R.T.: 0.000 min
Response: 94129676
Conc: 56.25 ng/ml

Instrument :
ECD_S
ClientSampleId :




#1 Dalapon

R.T.: 1.720 min
Delta R.T.: 0.000 min
Response: 371869843
Conc: 408.65 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 0.000 min
Exp R.T. : 4.810 min
Response: 0
Conc: N.D.



#2 3,5-DICHLOROBENZOIC ACID


R.T.: 4.515 min
Delta R.T.: 0.040 min
Response: 45787051
Conc: 53.21 ng/ml



#4 2,4-DCAA


R.T.: 5.453 min
Delta R.T.: 0.017 min
Response: 308274436
Conc: 225.24 ng/ml

Instrument :
ECD_S
ClientSampleId :




#4 2,4-DCAA

R.T.: 5.182 min
Delta R.T.: 0.013 min
Response: 159144025
Conc: 226.28 ng/ml




#5 DICAMBA

R.T.: 5.547 min
Delta R.T.: -0.039 min
Response: 42616082
Conc: 8.67 ng/ml



#5 DICAMBA


R.T.: 0.000 min
Exp R.T.: 5.294 min
Response: 0
Conc: N.D.



#7 MCPA


R.T.: 0.000 min
Exp R.T.: 5.867 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :




#7 MCPA

R.T.: 5.552 min
Delta R.T.: -0.003 min
Response: 23791627
Conc: 8.18 ug/ml




#9 2,4-D

R.T.: 0.000 min
Exp R.T.: 6.424 min
Response: 0
Conc: N.D.



#9 2,4-D


R.T.: 6.086 min
Delta R.T.: -0.009 min
Response: 11847480
Conc: 14.85 ng/ml



#10 Pentachlorophenol


R.T.: 0.000 min
Exp R.T.: 6.680 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :




#10 Pentachlorophenol

R.T.: 6.475 min
Delta R.T.: 0.017 min
Response: 9699009
Conc: 1.01 ng/ml




#11 2,4,5-TP (SILVEX)

R.T.: 0.000 min
Exp R.T.: 7.214 min
Response: 0
Conc: N.D.



#11 2,4,5-TP (SILVEX)


R.T.: 6.780 min
Delta R.T.: -0.041 min
Response: 58890156
Conc: 15.48 ng/ml



#12 2,4,5-T


R.T.: 7.422 min
Delta R.T.: -0.064 min
Response: 150033791
Conc: 19.63 ng/ml

Instrument :
ECD_S
ClientSampleId :




#12 2,4,5-T

R.T.: 0.000 min
Exp R.T. : 7.178 min
Response: 0
Conc: N.D.




#13 2,4-DB

R.T.: 0.000 min
Exp R.T. : 8.014 min
Response: 0
Conc: N.D.



#13 2,4-DB


R.T.: 7.680 min
Delta R.T.: 0.017 min
Response: 18768368
Conc: 40.04 ng/ml



#14 DINOSEB

R.T.: 0.000 min
Exp R.T.: 9.095 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :



#14 DINOSEB

R.T.: 8.009 min
Delta R.T.: 0.038 min
Response: 14563877
Conc: 5.19 ng/ml