

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS102422\
 Data File : PS020853.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Oct 2022 15:03
 Operator : AR\AJ
 Sample : N5186-01MS
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 TEXAS-SP-1019MS

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 10/25/2022
 Supervised By : Ankita Jodhani 10/25/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 24 18:37:02 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS100622.M
 Quant Title : 8080.M
 QLast Update : Fri Oct 07 06:09:52 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.418	6.857	733.7E6	492.1E6	388.937	367.515
Target Compounds							
1) T	Dalapon	2.510	2.578	303.7E6	496.6E6	108.286m	286.081m#
2) T	3,5-DICHL...	5.784	6.025	900.7E6	594.1E6	299.772	296.361
3) T	4-Nitroph...	6.252	6.462	359.8E6	186.7E6	266.198	226.483
5) T	DICAMBA	6.565	7.025	2616.2E6	1788.7E6	329.069	303.909
6) T	MCPD	6.720	7.121	228.0E6	113.5E6	38.807	25.895 #
7) T	MCPA	6.846	7.332	256.7E6	175.4E6	30.132	28.971
8) T	DICHLORPROP	7.166	7.653	604.3E6	422.9E6	301.234	291.558
9) T	2,4-D	7.365	7.942	790.5E6	553.5E6	324.215	312.271
10) T	Pentachlo...	7.602	8.379	7865.7E6	5068.2E6	286.466	250.904
11) T	2,4,5-TP ...	8.131	8.741	3418.1E6	2383.5E6	303.186	281.314
12) T	2,4,5-T	8.392	9.116	3744.8E6	2288.6E6	336.923	281.272
13) T	2,4-DB	8.900	9.626	595.1E6	340.3E6	260.533	303.777
14) T	DINOSEB	9.962	9.956	1493.5E6	1106.1E6	182.374	182.532
15) T	Picloram	9.784	10.925	6469.4E6	3768.0E6	388.465	319.655
16) T	DCPA	10.228	10.889	6331.4E6	3651.8E6	455.119	355.451

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS102422\
Data File : PS020853.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Oct 2022 15:03
Operator : AR\AJ
Sample : N5186-01MS
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
TEXAS-SP-1019MS

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 10/25/2022
Supervised By : Ankita Jodhani 10/25/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 24 18:37:02 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS100622.M
Quant Title : 8080.M
QLast Update : Fri Oct 07 06:09:52 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

