

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS102519\
 Data File : PS007336.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Oct 2019 11:23
 Operator : SM\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Ankita
 10/28/2019 10:55:52 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 25 12:03:26 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS101419.M
 Quant Title : 8080.M
 QLast Update : Mon Oct 14 13:54:23 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.461	5.189	946.7E6	479.6E6	691.717	681.893
Target Compounds							
1) T	Dalapon	2.042	1.752	1084.9E6	641.6E6	648.351	705.088m
2) T	3,5-DICHL...	4.836	4.497	1118.3E6	567.1E6	665.506	659.036
3) T	4-Nitroph...	5.314	4.879	374.4E6	235.8E6	623.580	597.086
5) T	DICAMBA	5.611	5.314	3288.8E6	1705.0E6	669.350	655.257
6) T	MCPD	5.763	5.398	238.1E6	123.6E6	67.565	65.144m
7) T	MCPA	5.894	5.573	368.6E6	187.8E6	71.461	64.606
8) T	DICHLORPROP	6.233	5.841	826.0E6	430.6E6	704.607	661.253
9) T	2,4-D	6.445	6.109	906.9E6	529.8E6	735.912	664.020
10) T	Pentachlo...	6.705	6.472	12647.0E6	6515.2E6	708.428	677.509
11) T	2,4,5-TP ...	7.237	6.834	4999.9E6	2619.0E6	717.384	688.562
12) T	2,4,5-T	7.507	7.192	5293.5E6	2521.8E6	692.437	719.758
13) T	2,4-DB	8.033	7.677	725.2E6	338.2E6	760.554	721.512
14) T	DINOSEB	9.118	7.986	3796.5E6	1908.0E6	652.413	679.627
15) T	Picloram	8.948	8.920	7498.9E6	3877.2E6	679.412	631.442
16) T	DCPA	9.392	8.879	6445.4E6	3504.2E6	719.812	698.793

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS102519\
Data File : PS007336.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Oct 2019 11:23
Operator : SM\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
HSTDCCC750

Manual Integrations
APPROVED

Ankita
10/28/2019 10:55:52 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 25 12:03:26 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS101419.M
Quant Title : 8080.M
QLast Update : Mon Oct 14 13:54:23 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

