

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS102521\
 Data File : PS017187.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Oct 2021 20:14
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

mohammad
 10/26/2021 10:35:09 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 26 02:04:56 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS101821.M
 Quant Title : 8080.M
 QLast Update : Tue Oct 19 01:21:46 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.151	6.737	1117.3E6	439.5E6	731.014m	751.956
Target Compounds							
1) T	Dalapon	2.989	2.493	1589.9E6	483.7E6	732.203	747.577
2) T	3,5-DICHL...	6.376	5.935	1483.5E6	615.6E6	687.207	708.767
3) T	4-Nitroph...	6.981	6.362	581.9E6	222.8E6	667.218	691.194
5) T	DICAMBA	7.341	6.897	4243.0E6	1831.7E6	708.381	719.437
6) T	MCPD	7.511	6.995	324.5E6	146.1E6	72.578	72.552
7) T	MCPA	7.660	7.206	446.6E6	207.0E6	72.394	69.923
8) T	DICHLORPROP	8.032	7.515	1056.3E6	459.5E6	704.489	714.492
9) T	2,4-D	8.261	7.805	1260.1E6	545.3E6	711.451	724.510
10) T	Pentachlo...	8.566	8.230	14817.3E6	5967.9E6	714.153	730.946
11) T	2,4,5-TP ...	9.116	8.590	5851.2E6	2536.1E6	719.462	735.717
12) T	2,4,5-T	9.400	8.967	5916.6E6	2394.6E6	736.618	755.510
13) T	2,4-DB	9.949	9.471	1165.7E6	300.1E6	774.888	723.571
14) T	DINOSEB	11.102	9.797	4393.7E6	1783.3E6	775.609	808.168
15) T	Picloram	10.923	10.775	8369.9E6	3078.6E6	783.597	803.860
16) T	DCPA	11.370	10.726	7549.1E6	3159.5E6	741.254	740.400

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS102521\
Data File : PS017187.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Oct 2021 20:14
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

mohammad
10/26/2021 10:35:09 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 26 02:04:56 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS101821.M
Quant Title : 8080.M
QLast Update : Tue Oct 19 01:21:46 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

