

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS102523\
 Data File : PS024706.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Oct 2023 10:32
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 10/26/2023
 Supervised By : Ankita Jodhani 10/26/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 25 12:46:14 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS100323.M
 Quant Title : 8080.M
 QLast Update : Tue Oct 03 14:54:54 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.357	6.777	1516.9E6	499.9E6	737.300	723.182
Target Compounds							
1) T	Dalapon	2.477	2.538	1631.7E6	699.9E6	660.466	625.417
2) T	3,5-DICHL...	5.732	5.964	2151.5E6	744.6E6	685.887	685.208
3) T	4-Nitroph...	6.195	6.392	895.0E6	297.4E6	661.169	651.949
5) T	DICAMBA	6.498	6.938	6361.0E6	2064.0E6	709.088	696.987m
6) T	MCPD	6.650	7.035	457.7E6	138.9E6	72.244	70.487m
7) T	MCPA	6.772	7.241	630.1E6	210.7E6	70.702	65.759
8) T	DICHLORPROP	7.085	7.552	1614.9E6	501.3E6	686.970	666.597
9) T	2,4-D	7.283	7.836	1907.3E6	613.1E6	690.293	666.121
10) T	Pentachlo...	7.512	8.260	23693.8E6	7619.9E6	728.707	699.316
11) T	2,4,5-TP ...	8.033	8.618	9855.2E6	2926.9E6	711.307	690.133
12) T	2,4,5-T	8.289	8.987	9268.1E6	2849.0E6	709.370	690.902
13) T	2,4-DB	8.793	9.490	1556.6E6	358.0E6	761.959	702.568
14) T	DINOSEB	9.846	9.818	7074.5E6	2082.8E6	692.045	671.657
15) T	Picloram	9.673	10.769	12121.2E6	4191.0E6	695.171	674.408
16) T	DCPA	10.112	10.739	12115.9E6	3717.2E6	711.747	712.457m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS102523\
Data File : PS024706.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Oct 2023 10:32
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 10/26/2023
Supervised By : Ankita Jodhani 10/26/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 25 12:46:14 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS100323.M
Quant Title : 8080.M
QLast Update : Tue Oct 03 14:54:54 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

