

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS102621\
 Data File : PS017230.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Oct 2021 19:46
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 27 01:12:07 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS101821.M
 Quant Title : 8080.M
 QLast Update : Tue Oct 19 01:21:46 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.152	6.737	1101.2E6	418.1E6	720.483	715.391
Target Compounds							
1) T	Dalapon	2.987	2.492	1482.9E6	452.5E6	682.903	699.309
2) T	3,5-DICHL...	6.375	5.935	1412.8E6	586.5E6	654.468	675.329
3) T	4-Nitroph...	6.981	6.362	557.5E6	213.7E6	639.265	662.893
5) T	DICAMBA	7.341	6.899	4024.7E6	1742.9E6	671.927	684.561
6) T	MCPD	7.511	6.995	305.4E6	138.6E6	68.312	68.845
7) T	MCPA	7.659	7.206	424.1E6	195.3E6	68.742	65.970
8) T	DICHLORPROP	8.033	7.515	1007.6E6	440.6E6	671.993	685.156
9) T	2,4-D	8.261	7.806	1204.1E6	524.3E6	679.830	696.711
10) T	Pentachlo...	8.566	8.230	14222.6E6	5725.5E6	685.489	701.257
11) T	2,4,5-TP ...	9.118	8.591	5582.6E6	2425.0E6	686.434	703.480
12) T	2,4,5-T	9.401	8.968	5671.5E6	2300.8E6	706.097	725.925
13) T	2,4-DB	9.950	9.473	1125.0E6	289.2E6	747.834	697.335
14) T	DINOSEB	11.104	9.799	4207.6E6	1707.1E6	742.766	773.625
15) T	Picloram	10.925	10.777	8113.7E6	2990.7E6	759.604	780.908
16) T	DCPA	11.372	10.729	7216.1E6	3030.1E6	708.555	710.088

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS102621\
Data File : PS017230.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Oct 2021 19:46
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 27 01:12:07 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS101821.M
Quant Title : 8080.M
QLast Update : Tue Oct 19 01:21:46 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

